

ENCLOSURE

NOTICE OF VIOLATION

Veterans Administration Medical Center

License No. 01-03042-02

The following violation was identified during an inspection conducted on June 26, 1985. The Severity Level was assigned in accordance with the NRC Enforcement Policy (10 CFR Part 2, Appendix C).

License Condition 18 requires that licensed material be used in accordance with statements and procedures contained in your application dated July 31, 1979. One procedure in your application requires your dose calibrator to be tested for linearity at quarterly intervals.

Contrary to the above, the dose calibrator in Nuclear Medicine had not been tested for linearity since August 1984.

This is a Severity Level IV violation (Supplement VI).

Pursuant to 10 CFR 2.201, you are required to submit to this office within 30 days of the date of this Notice, a written statement or explanation in reply, including: (1) admission or denial of the alleged violations; (2) the reasons for the violations if admitted; (3) the corrective steps which have been taken and the results achieved; (4) corrective steps which will be taken to avoid further violations; and (5) the date when full compliance will be achieved.

Security or safeguards information should be submitted as an enclosure to facilitate withholding it from public disclosure as required by 10 CFR 2.790(d) or 10 CFR 73.21.

Date: JUL 11 1985