

# Petricca Industries

## Integrated Construction

P.O.Box 1145—444 Merrill Rd.  
Pittsfield, MA 01201  
(413) 442-6926



July 5, 1985

United States  
Nuclear Regulatory Commission  
Region I  
631 Park Avenue  
King of Prussia, Pennsylvania 19406

License No. 20-20591-01

Gentlemen:

We deeply regret the results of the routine safety inspection conducted by Mr. C.T. Oberg on March 25, 1985. It is our intention to be in full compliance with NRC requirements and to maintain them. The following listing of corrective and preventive steps are for condition in which we violated NRC regulations.

- A. License condition No. 12. Copies of certificate of training for each user have been obtained and are available.
- B. 10 CFR 20.207, 10 CFR 20.207, 20.2 (a)(17) The troxler gauge is now stored in a locked cabinet and is under constant surveillance. It is company policy to store the troxler gauge in its locked cabinet at the end of each day.
- C. License condition No. 15. A physical inventory on all sealed sources has been completed and records are being stored.
- D. License condition No. 14. A leakage test was performed the week after your inspection. A troxler gauge log book is being used to remind us when test and other items are to be performed.
- E. 10 CFR 20.401 (a) Exposure records are currently being complied. In the future all information required by the NRC-5 form will be complied with.
- F. 10 CFR 19.11 (a) & (b) 10 CFR 19.11 (d) Current copies of license conditions and documents are posted at locations of high traffic. We have set up a troxler bulletin board for our licensed people.

8508080425 850729  
REG1 LIC30  
20-20591-01 PDR

# Petricca Industries



United States  
Nuclear Regulatory Commission  
Page 2  
July 5, 1985

We hope that these corrections are to your satisfaction and we plan to be in full compliance by July 31, 1985.

If you have any questions please feel free to contact me.

Very truly yours,

PETRICCA CONSTRUCTION COMPANY

(A Subsidiary Division of Petricca Industries, Inc.)

A handwritten signature in cursive script that reads "Peter F. Libardi".

Peter F. Libardi,  
General Superintendent

PFL/mjm