

## Material Licensee Temporary Exemptions During the COVID-19 Public Health Emergency

05/07/2020

This table provides a list of 10 CFR Part 30 requirements for the which the NRC may consider expedited requests for temporary exemption. Licensees may seek a temporary exemption from these requirements to address the challenges licensees may face during the COVID-19 Public Health Emergency (PHE). This table may be updated as the NRC identifies additional requirements for which the NRC may consider expedited requests for temporary exemption.

Reg.	Description of Exemption
10 CFR 30.34 Terms and conditions of licenses	
<a href="#">30.34</a>	The license condition requiring that a physical inventory be performed every 6 months
<a href="#">30.34</a>	The requirement in License Condition <b>[number]</b> to perform leak testing every <b>[X]</b> months or at the frequency specified in the SSD <i>Note: The leak testing requirements specified in the SSD range from 6 months to 3 years.</i>
<a href="#">30.34</a>	The requirement in License Condition <b>[number]</b> to perform surveys
<a href="#">30.34</a>	The requirement in License Condition <b>[number]</b> to exchange dosimetry
<a href="#">30.34</a>	The requirement in License Condition <b>[number]</b> to provide radiation safety refresher training at intervals not to exceed 12 months
<a href="#">30.34</a>	The requirement in License Condition <b>[number]</b> to perform instrument calibration
<a href="#">30.34</a>	The requirement in License Condition <b>[number]</b> to hold Radiation Safety Committee meetings at intervals not to exceed <b>[X]</b> months
<a href="#">30.34</a>	The requirement in License Condition <b>[number]</b> to complete an annual program review at intervals not to exceed 12 months