



Winthrop Hospital, Inc.

40 Lincoln Street · Winthrop, MA 02152

846-2600

John E. Glen, Ph.D
Chief Nuclear Materials & Safeguard Branch
US Nuclear Regulatory Commission
631 Park Avenue
King of Prussia, PA 19406


Dear Dr. Glen:

In reference to your June 17, 1985 letter:

1. Linearity tests will be performed on a quarterly basis using the calicheck procedure system provided by Calcorp, Cleveland, Ohio and performed by Neal Gaeta as per his NRC license 20-20743-01.
2. At this time we wish to drop from consideration our request for Group III license amendative.

When the appropriate documentation has been gathered, resubmission for Group III use will be made.

Sincerely,



Gerard Nocton
Executive Director

1985 JUL 15 PM 2:03
RECEIVED - REGION 1

03603

"OFFICIAL RECORD COPY"

8508120155 850726
REG1 LIC30
20-18110-01 PDR

ML10

JUL 15 1985