

MATERIALS LICENSE

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter 1, Parts 30, 31, 32, 33, 34, 35, 40 and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

Licensee	
1. Oldover Corporation	3. License number 45-23547-01
2. P.O. Box 228 Ashland, Virginia 23005	4. Expiration date July 31, 1990
	5. Docket or Reference No. 030-28768
6. Byproduct, source, and/or special nuclear material	7. Chemical and/or physical form
A. Nickel 63	A. Foil source (New England Nuclear Model NER004)
	8. Maximum amount that licensee may possess at any one time under this license
	A. Not to exceed 10 millicuries per foil per device
9. Authorized use	
A. Foil source in Shimadzu Scientific Model GC-3AIE detector cell for use in gas chromatograph for sample analysis.	

★ ★ ★ CONDITIONS ★ ★ ★

10. Licensed material shall be used only at the licensee's facilities located on U.S. Route 1, 3 miles south of Ashland, Virginia.
11. The licensee shall comply with the provisions of Title 10, Chapter 1, Code of Federal Regulations, Part 19, "Notices, Instructions and Reports to Workers; Inspections" and Part 20, "Standards for Protection Against Radiation."
12. Licensed material shall be used by, or under the supervision of, Jonathan Benz, Donald S. Grigsby, or Louis R. Lawson, Jr.
13. Foil sources containing licensed material shall not be opened or removed from detector cells by the licensee.

MATERIALS LICENSE
SUPPLEMENTARY SHEET

License number:

45-23547-01

Docket or Reference number:

030-28768

(con't)

CONDITIONS

14. In lieu of using the conventional radiation caution colors (magenta or purple on yellow background) as provided in Section 20.203(a)(1), Title 10, Code of Federal Regulations, Part 20, the licensee is hereby authorized to label detector cells and cell baths, containing licensed material and used in gas chromatography devices, with conspicuously etched or stamped radiation caution symbols without a color requirement.
15. A. Each chromatograph detector containing Nickel 63 shall be tested for leakage and/or contamination at intervals not to exceed six months. In the absence of a certificate from a transferor indicating that a test has been made within six months prior to the transfer, a detector received from another person shall not be put into use until tested.
- B. The test shall be capable of detecting the presence of 0.005 microcurie of radioactive material on the test sample. The test sample shall be taken from the surfaces of the device in which the foil is mounted or stored on which one might expect contamination to accumulate. Records of leak test results shall be kept in units of microcuries and maintained for inspection by the Commission.
- C. If the test reveals the presence of 0.005 microcurie or more of removable contamination, the licensee shall immediately withdraw the foil from use and shall cause it to be decontaminated and repaired or to be disposed of in accordance with Commission regulations. A report shall be filed within five (5) days of the test with the U.S. Nuclear Regulatory Commission, Region II, Division of Radiation Safety and Safeguards, Nuclear Material Safety Section, 101 Marietta Street, Suite 2900, Atlanta, Georgia 30323, describing the equipment involved, the test results, and the corrective action taken.
- D. The licensee is authorized to collect leak test samples in accordance with the procedures described in the licensee's application dated June 27, 1985 for analysis by Nuclear Radiation Development Corporation. Alternatively, leak test samples may be collected and/or analyzed by other persons specifically authorized by the Commission or an Agreement State to perform such services.
16. The licensee shall conduct a physical inventory every six (6) months to account for all foil sources received and possessed under the license. The records of the inventories shall be maintained for two (2) years from the date of the inventory for inspection by the Commission and shall include the quantities and kinds of byproduct material, manufacturer's name and model numbers, location of detector cells and the date of the inventory.

MATERIALS LICENSE
SUPPLEMENTARY SHEETLicense number
45-23547-01Docket or Reference number
030-28768

(con't)

CONDITIONS

17. Except as specifically provided otherwise by this license, the licensee shall possess and use licensed material described in Items 6, 7, and 8 of this license in accordance with statements, representations, and procedures contained in application dated June 27, 1985. The Nuclear Regulatory Commission's regulations shall govern the licensee's statements in applications or letters, unless the statements are more restrictive than the regulations.



FOR THE U.S. NUCLEAR REGULATORY COMMISSION

EARL G. WRIGHT

Date AUG 01 1985

By

Earl G. WrightRegion II, Nuclear Materials
Safety Section
101 Marietta Street, Suite 2900
Atlanta, GA 30323

7/24/85

TELEPHONE OR VERBAL CONVERSATION RECORD

TIME

☐ A.M.

☐ P.M.

☐ INCOMING CALL

☒ OUTGOING CALL

☐ VISIT

PERSON CALLING

E Wright

OFFICE/ADDRESS

PHONE NUMBER

EXTENSION

PERSON CALLED

Louis R Lawson, Jr

OFFICE/ADDRESS

& Oldover Corp

PHONE NUMBER

EXTENSION

804-798-7981

CONVERSATION -

SUBJECT

Re: control 19082 license application

SUMMARY

Called to verify descriptive address
of UAE.

US Route 1 North
3 miles south of Ashland, Va.

Earl Wright

REFERRED TO:

ACTION REQUESTED

☐ ADVISE ME OF
ACTION TAKEN.

INITIALS

DATE

ACTION TAKEN

INITIALS

DATE