


THE WASHINGTON HOSPITAL CENTER



June 3, 1985

U.S. Nuclear Regulatory Commission
Region I
631 Park Ave
King of Prussia, Pa 19406

Dear Sirs:


Since the Amendment to our license 08-03604-03 (letter of April 17, 1985) has not yet been processed I would like to request that the following information be considered as part of the Amendment request.

Please amend license 08-03604-03 to include the following radioactive material source:

Gd-153 sealed source as GdO₂ manufactured by Gulf Nuclear Model GD-1 containing 1000 millicuries of activity. A total possession limit of 1300 mCi is requested to include the summed activity of a newly received source and the decayed source to be returned to the source manufacturer.

The above source will be used in a Lunar DP3 bone mineral analyzer. The NRC device registration number is NR-430-D-101-S. The device will be located and used in the Nuclear Medicine Dept. Training is to be provided by the installation personnel and the RSO will be present during the training sessions and for instructions on source replacement and wipe testing.

Sincerely yours,


Kenneth D. Williams
Radiation Safety Office

KDW/pmr

03751

"OFFICIAL RECORD COPY"

ML10

8508120228 850726
REG1 LIC30
08-03604-03 PDR