

JUN 25 1985

Docket No. : 030-17576
License No.: 02-19418-01
Control No.: 19007

U.S. Department of the Interior
Bureau of Indian Affairs
Colorado River Agency
Route 1, Box 9-C
Parker, Arizona 85344

Att: Mr. Murdock
Superintendent

This is in reference to your application dated May 17, 1985 to renew your byproduct material license. In order to complete our review, we need the following additional information:

1. Please submit further information on the location of your Poston, Arizona facility. If there is no street address, you should submit a map which describes the location. Also, describe how the gauges are secured to ensure against unauthorized removal or use as required by 10 CFR 20.207. Refer to Item 9. of the enclosed guide.
2. Please note that Item 8., page 6 of the enclosed guide states that each employee should complete the device manufacturer's training program before being permitted to use the device. Otherwise, the employee must use the gauge under the supervision and in the physical presence of another individual who has completed the device manufacturer's training course. You should submit such a commitment.
3. State the type of personnel monitoring equipment you will use (film badge or thermoluminescent dosimeter). Film badges should be exchanged monthly. Thermoluminescent dosimeters (TLD) may be exchanged quarterly. Refer to Item 10.1 on pages 7 and 8 of the enclosed guide.
4. You should specify the name, address, and license number of the consultant or commercial organization which will conduct the analysis of your leak tests. Refer to Item 10.3 of the enclosed guide.
5. You should submit a description of your operating and emergency procedures. Refer to Item 10.6 of the enclosed guide.

8510230299 850821
REG5 LIC30
02-19418-01 PDR

OFFICE	RV <i>BAR</i>						
SURNAME	Riedlinger:fr						
DATE	6/25/85						

U.S. Dept. of the Interior

-2-

JUN 25 1985

We will continue the review of your renewal request upon receipt of this information. You should submit your response within thirty (30) days from the date of this notice. Please reply in duplicate, and refer to Mail Control No. 19007.

Sincerely,

Beth A. Riedlinger
Health Physicist (Licensing)
Nuclear Materials Safety Section

Enclosure: Licensing Guide for Portable Gauges