

"OFFICIAL RECORD COPY"

JUL 26 1985

Tennessee Valley Authority
ATTN: George Hoffmeister or G. T. Jones
National Fertilizer Development Center
Muscle Shoals, AL 35660

Gentlemen:

SUBJECT: REQUEST FOR MORE INFORMATION ABOUT YOUR APPLICATION
(Reference: 19031; 030-03571)

This refers to your letter application dated May 31, 1985 for amendment of License No. 01-06113-03.

Please provide the following supplementary information in support of your request.

1. A clear and specific statement of procedures for installation, initial radiation survey and leak testing of the Nuclear Research level gauge. If this will be done by someone other than the device manufacturer please provide specific information as specified on page 8, 9 and 10 of the enclosed NRC guide.
2. A description of procedures that will be instituted to prevent exposure of personnel during maintenance and clean up associated with the device and equipment it is used with. (For example, lockout procedures to prevent personnel from entering bins, tanks or vats when high radiation levels are present.) Please refer to paragraph 10.5, page 11 of the enclosed guide.
3. A description of the intended use of the device to include information to show that: (a) personnel access is limited to prevent unnecessary exposure during operation and maintenance and (b) an analysis of potential environmental factors which may effect the safe operation of the device in its intended use (for example elevated temperatures, corrosive conditions, excessive vibration or fire or explosion) please refer to Item 9, pages 7 and 8 and provide the information specified in subitems 1-6.

To facilitate the issuance of your license we ask that you provide the requested information within 30 days of the receipt of this letter. Otherwise, we shall assume that you have abandoned your application.

If you have questions about this letter, please call me. My telephone number is (404) 221-5565.

Sincerely,

8510240008 850917
REG2 LIC30
01-06113-03 PDR

Earl G. Wright
Senior License Reviewer

Enclosure:
Regulatory Guide for Nonportable
Gauges dated January 1985

RII

Right:sa
7/25/85

76 JPP
7/26/85

7/2/85

TELEPHONE OR VERBAL CONVERSATION RECORD

TIME

☒ A.M.

☐ P.M.

☐ INCOMING CALL

☒ OUTGOING CALL

☐ VISIT

PERSON CALLING

Earl Wright

OFFICE/ADDRESS

PHONE NUMBER

EXTENSION

PERSON CALLED

Steve Baggett

OFFICE/ADDRESS

FC/NM55

PHONE NUMBER

EXTENSION

427-9005

CONVERSATION -

SUBJECT

Re: Nuclear Research Model SH 352-50

SUMMARY

Steve said This is currently on The
NRC "G" license + a new sheet
Will be out in about 1 month.

OK TO license-

gfw

REFERRED TO:

ACTION REQUESTED

☐ ADVISE ME OF
ACTION TAKEN.

INITIALS

DATE

ACTION TAKEN

INITIALS

DATE