

AUG 21 1985

Docket No. : 030-01209  
License No.: 02-10072-01  
Control No.: 16928

V. A. Medical Center  
7th Street and Indian School Road  
Phoenix, Arizona 85012

Attention: Mr. Ray L. Bourne  
Medical Center Director

Gentlemen:

Enclosed is the NRC license renewal which you requested.

Please review the enclosed document carefully and be sure that you understand all conditions. If there are any errors or questions, please notify us so that we can provide appropriate corrections and answers.

Since you are authorized to use I-131 for treatment of thyroid carcinoma, you are expected to have a pancake or end-window survey meter, as stated in your letter dated July 8, 1985. This instrument should be used for surveys of the patient's room. This matter will be reviewed during the next routine inspection.

Any future correspondence relating to your license should specifically reference your license and docket numbers to expedite your inquiry.

Sincerely,

Beth A. Riedlinger  
Health Physicist (Licensing)  
Nuclear Materials Safety Section

Enclosure: Attachment A  
Amendment No. 34 to License No. 02-10072-01

8510240049 850821  
REG5 LIC30  
02-10072-01 PDR

OFFICE	B&R						
SURNAME							
DATE	8/21/85						