

MATERIALS LICENSE

Amendment No. 02

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 40 and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

Licensee	
1. Allgeier, Martin and Associates, Inc.	In accordance with letter dated May 22, 1985
2. P.O. Box 2627 Joplin, MO 64803	3. License number 24-18942-01 is amended in its entirety to read as follows:
	4. Expiration date August 31, 1990
	5. Docket or Reference No. 030-17383
6. Byproduct, source, and/or special nuclear material	7. Chemical and/or physical form
A. Cesium-137	A. Sealed source (Campbell Pacific Nuclear Model No. CPN-131)
B. Americium-241	B. Sealed source (Campbell Pacific Nuclear Model No. CPN-131)
	8. Maximum amount that licensee may possess at any one time under this license
	A. No single source to exceed 10 millicuries
	B. No single source to exceed 50 millicuries
9. Authorized Use	
A. and B. To be used in Campbell Pacific Nuclear MC Series moisture/density gauge to measure moisture/density of construction materials.	

CONDITIONS

10. Licensed material may be stored and used at the licensee's facilities located at 2820 Rangeline Road, Joplin, Missouri and may be used at temporary job sites throughout the States of Missouri and Oklahoma.
11. The licensee shall comply with the provisions of Title 10, Chapter 1, Code of Federal Regulations, Part 19, "Notices, Instructions and Reports to Workers; Inspections" and Part 20, "Standards for Protection Against Radiation."
12. Licensed material shall be used by, or under the supervision and in the physical presence of, individuals who have satisfactorily completed the device manufacturer's training program for gauge users and have been designated by the licensee's Radiation Protection Officer. The licensee shall maintain records of the individuals who have been designated as authorized users.

8508120706 850801
REG3 LIC30
24-18942-01 PDR

1/1 ML 30 COPY 2
8/1/85 cm

MATERIALS LICENSE
SUPPLEMENTARY SHEET

License number

24-18942-01

Docket or Reference number

030-17383

Amendment No. 02

13. A. Each sealed source containing licensed material, other than Hydrogen 3, with a half-life greater than thirty days and in any form other than gas shall be tested for leakage and/or contamination at intervals not to exceed six months. In the absence of a certificate from a transferor indicating that a test has been made within six months prior to the transfer, a sealed source received from another person shall not be put into use until tested.
- B. The test shall be capable of detecting the presence of 0.005 microcurie of radioactive material on the test sample. The test sample shall be taken from the sealed source or from the surfaces of the device in which the sealed source is permanently mounted or stored on which one might expect contamination to accumulate. Records of leak test results shall be kept in units of microcuries and maintained for inspection by the Commission.
- C. If the test reveals the presence of 0.005 microcurie or more of removable contamination, the licensee shall immediately withdraw the sealed source from use and shall cause it to be decontaminated and repaired or to be disposed of in accordance with Commission regulations. A report shall be filed within five (5) days of the test with the U. S. Nuclear Regulatory Commission, Region III, 799 Roosevelt Road, Glen Ellyn, Illinois 60137, describing the equipment involved, the test results, and the corrective action taken.
- D. The licensee is authorized to collect leak test samples in accordance with the procedures described in the licensee's letter dated July 23, 1980 for analysis by Campbell Pacific Nuclear. Alternatively, leak test samples may be collected and/or analyzed by other persons specifically authorized by the Commission or an Agreement State to perform such services.
14. Sealed sources containing licensed material shall not be opened or removed from the moisture/density gauge by the licensee.
15. When performing tests at temporary job sites, the authorized user shall not leave the moisture/density gauge unattended. Upon completion of tests the device shall be locked in the licensee's vehicle or a secure building to prevent unauthorized use, loss or theft.
16. The licensee shall conduct a physical inventory every six (6) months to account for all gauges received and possessed under the license. The records of the inventories shall be maintained for two (2) years from the date of the inventory for inspection by the Commission, and shall include the quantities and kinds of byproduct material, manufacturer's name and model numbers, location of the gauges and the date of the inventory.
17. The licensee may transport licensed material or deliver licensed material to a carrier for transport in accordance with the provisions of Title 10, Code of Federal Regulations, Part 71, "Packaging of Radioactive Material for Transport and Transportation of Radioactive Material Under Certain Conditions."

COPY

MATERIALS LICENSE
SUPPLEMENTARY SHEET

License number

24-18942-01

Docket or Reference number

030-17383

Amendment No. 02

18. Maintenance and repair on the device that involves removal of the source or its placement in an unshielded position shall be performed by the device manufacturer or by other persons specifically authorized by the Commission or an Agreement State to perform such services.
19. Except as specifically provided otherwise by this license, the licensee shall possess and use licensed material described in Items 6, 7, and 8 of this license in accordance with statements, representations, and procedures contained in letters dated May 19, 1980 and July 23, 1980 with attachments. The Nuclear Regulatory Commission's regulations shall govern the licensee's statements in applications or letters, unless the statements are more restrictive than the regulations.



For the U.S. Nuclear Regulatory Commission

Original Signed

By Evelyn R. Matson

Materials Licensing Section, Region III

Date August 1, 1985

COPY