

JUL 31 1985

The Bionetics Corporation
ATTN: Lory B. Littlefield, Ph.D.
Project Manager
16 Triangle Park Drive
Cincinnati, OH 45246

Gentlemen:

Enclosed is Amendment No. 01 renewing your NRC License No. 34-18805-01 in accordance with your request.

Please review the enclosed document carefully and be sure that you understand all conditions. You must conduct your program involving radioactive materials in accordance with the conditions of your NRC license, representations made in your license application, and NRC regulations. In particular, note that you must:

1. Operate in accordance with NRC regulations 10 CFR Part 19, "Notices, Instructions and Reports to Workers; Inspections," 10 CFR Part 20, "Standards for Protection Against Radiation," and other applicable regulations.
2. Possess radioactive material only in the quantity and form indicated in your license.
3. Use radioactive material only for the purpose(s) indicated in your license.
4. Notify NRC in writing of any change in mailing address.
5. Request and obtain appropriate amendment if you plan to change ownership of your organization, change locations of radioactive material, or make any other changes in your facility or program which are contrary to your license conditions or representations made in your license application and any supplemental correspondence with NRC. Any amendment request should be accompanied by the appropriate fee specified in 10 CFR Part 170.
6. Submit a complete renewal application with proper fee or termination request at least 30 days before the expiration date on your license. You will receive a reminder notice approximately 90 days before the expiration date. Possession of radioactive material after your license expires is a violation of NRC regulations.
7. Request termination of your license if you plan to permanently discontinue activities involving radioactive material prior to your expiration date.

8508120728 850731
REG3 LIC30
34-18805-01 PDR

You will be periodically inspected by NRC. Failure to conduct your program in accordance with NRC regulations, license conditions and representations in your license application will result in enforcement action against you in accordance with the General Policy and Procedures for NRC Enforcement Actions, 10 CFR Part 2, Appendix C.

If you have any questions or require clarification of any of the above stated information, contact us at (312) 790-5625.

Sincerely,

Original Signed By
Bruce S. Mallett
Materials Licensing Section

Enclosure(s): Amendment No. 01

RIII

Mallett/cm
06/27/85

CONVERSATION RECORD

TIME

10:00 AM

DATE

18 JULY 85

TYPE

☐ VISIT☐ CONFERENCE☒ TELEPHONE☐ INCOMING☒ OUTGOING

ROUTING

NAME/SYMBOL

INT

Location of Visit/Conference:

NAME OF PERSON(S) CONTACTED OR IN CONTACT
WITH YOUORGANIZATION (Office, dept., bureau,
etc.)

TELEPHONE NO.

LORY LITTLEFIELD, Ph.D. BIONETICS CORP. 771-0448

SUBJECT

RENEWAL OF 34-18805-01

SUMMARY

- ① Model # of Detector Cell - confirmed that Thacor Model 115500-0001 is the correct model of the Thacor cell.
- ② Leak Test - leak test procedure of application dated 9-11-79 is O.K. to reference as current procedure. Dr. Littlefield requested a 3-year frequency.
- ③ Model 222 GC belonged to EPA (Cincinnati). Bionetics only used it on old license. It was returned to EPA and replaced w/ Model 560. ACTION REQUIRED at the new location. No BPM remains at the old address (Chesterdale Road).

NAME OF PERSON DOCUMENTING CONVERSATION

JEFF MUELLER

SIGNATURE

Jeff Mueller

DATE

18 July 85

ACTION TAKEN

SIGNATURE

TITLE

DATE