

"OFFICIAL RECORD COPY"

AUG 26 1985

Automata, Inc.
ATTN: Mr. Christopher L. Schoebe
Radiation Safety Officer
11091 Sunset Hills Road
Reston, VA 22090

Gentlemen:

SUBJECT: REQUEST FOR ADDITIONAL INFORMATION CONCERNING A MATERIAL LICENSE
APPLICATION, (REFERENCE: 50664; 030-20735)

This refers to your application dated July 10, 1985, requesting amendment of license number 45-20697-01. Additional information is needed in support of your application.

Please provide the following information:

1. Sterling, Virginia, Facility. An attachment to your application shows a facility sketch, and we are unable to determine if this is your new facility in Sterling, Virginia, or your present facility in Reston, Virginia. Please identify the sketch and if it is your Sterling, Virginia, facility, it is only necessary for you to specify this.
2. Authorized Users. Your July 10, 1985, application lists three individual users who are not presently listed in Condition 12 of your license. Is it your intent to delete the names of the three users currently listed on your license, i.e., list only the names shown in your July 10, 1985, application? Please provide information clearly showing the names of your employees to be listed on your license.
3. Training of Individual Users. An attachment to your application lists the names of four employees who completed the Glenbrook Company training program in the use of Lixiscope devices. We need to know if it is your intent that each employee to be named as an individual user on your license will complete the Glenbrook training program.
4. Transportation. We have assumed that since you will now have two facilities in which the Lixiscope devices will be used that these devices will be transported, as needed, from one facility to another. Please provide brief information showing that you will meet all requirements of the enclosed 10 CFR Part 71, "Packaging of Radioactive Material for Transport and Transportation of Radioactive Material Under Certain Conditions."

Our review of your application will continue upon receipt of the above information. Please provide two copies of your reply and reference Mail Control Number 50664.

8510240614 850917
REG2 LIC30
45-20697-01 PDR

Automata, Inc.
Reston, VA

2


AUG 26 1985

If we do not receive a reply from you within 30 calendar days from the date of this letter, we shall assume that you do not wish to pursue your application.

Sincerely,

Paul R. Guinn
Senior License Reviewer
Nuclear Materials Safety Section

Enclosure:
10CFR Part 71

RII 
PGuinn:nw
8/26/85

RII 
JPPotter
8/26/85

CONVERSATION RECORD

TIME

9:45 AM

DATE

8/12/85

TYPE

☐ VISIT☐ CONFERENCE☒ TELEPHONE☐ INCOMING☒ OUTGOING

ROUTING

NAME/SYMBOL

INT

Location of Visit/Conference:

NAME OF PERSON(S) CONTACTED OR IN CONTACT WITH YOU

Christopher L. Scholbe

ORGANIZATION (Office, dept., bureau, etc.)

Autonala, Inc
Reston, VA

TELEPHONE NO.

703-
471-5110

SUBJECT

Amendment of Lic. # 45-20697-01
Control # 50664

SUMMARY

talked to Susan J. Hoyer (she was told to Scholbe)
 asked ~~Scholbe~~ ^{Hoyer} to provide answers to the
 following questions:

1. In the drawing submitted with the 7/10/85 application for the new location in Sterling Virginia?
2. Are the men listed for the 7/10/85 intended to be the only men? (Replace previous men?)
3. How transported between facilities in Reston, Virginia and Sterling Va.?
4. Are men complete the Glenbrook Co. Training Course?

ACTION REQUIRED

Wait for phone call from Hoyer - (may want answers in writing)

NAME OF PERSON DOCUMENTING CONVERSATION

PAUL R. GUINA

SIGNATURE

Paul R. Guina

DATE

8/12/85

ACTION TAKEN

SIGNATURE

TITLE

DATE