

MATERIALS LICENSE

Amendment No. 01

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 40 and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

Licensee		In accordance with application dated June 25, 1985	
1. Fairfax County Government (Lower Potomac Plant Laboratory)		3. License number 45-19479-01 is amended in its entirety to read as follows:	
2. Box 268 Lorton, Virginia 22079		4. Expiration date August 31, 1990	
		5. Docket or Reference No. 030-17756	
6. Byproduct, source, and/or special nuclear material	7. Chemical and/or physical form	8. Maximum amount that licensee may possess at any one time under this license	
A. Nickel 63	A. Foils in Perkin-Elmer Model 330-0119 detector cells	A. Not to exceed 15 millicuries per foil	
9. Authorized use			
A. For use in gas chromatographs for sample analysis.			

CONDITIONS

10. Licensed material shall be used only at licensee's facilities at 9399 Richmond Highway, Lorton, Virginia.
11. The licensee shall comply with the provisions of Title 10, Chapter 1, Code of Federal Regulations, Part 19, "Notices, Instructions and Reports to Workers; Inspections" and Part 20, "Standards for Protection Against Radiation."
12. Licensed material shall be used by, or under the supervision of, David Bushman.
13. In lieu of using the conventional radiation caution colors (magenta or purple on yellow background) as provided in Section 20.203(a)(1), Title 10, Code of Federal Regulations, Part 20, the licensee is hereby authorized to label detector cells and cell baths, containing licensed material and used in gas chromatography devices, with conspicuously etched or stamped radiation caution symbols without a color requirement.

8508130203 850801
REG2 LIC30
45-19479-01 PDR

MLD

MATERIALS LICENSE
SUPPLEMENTARY SHEET

License number

45-19479-01

Docket or Reference number

030-17756

Amendment No. 01

(cont'd)

CONDITIONS

14. A. Each chromatograph detector containing Nickel 63 shall be tested for leakage and/or contamination at intervals not to exceed six months. In the absence of a certificate from a transferor indicating that a test has been made within six months prior to the transfer, a detector received from another person shall not be put into use until tested.
- B. The test shall be capable of detecting the presence of 0.005 microcurie of radioactive material on the test sample. The test sample shall be taken from the surfaces of the device in which the foil is mounted or stored on which one might expect contamination to accumulate. Records of leak test results shall be kept in units of microcuries and maintained for inspection by the Commission.
- C. If the test reveals the presence of 0.005 microcurie or more of removable contamination, the licensee shall immediately withdraw the foil from use and shall cause it to be decontaminated and repaired or to be disposed of in accordance with Commission regulations. A report shall be filed within five (5) days of the test with the U.S. Nuclear Regulatory Commission, Region II, Division of Radiation Safety and Safeguards, Nuclear Material Safety Section, 101 Marietta Street, Suite 2900, Atlanta, Georgia 30323, describing the equipment involved, the test results, and the corrective action taken.
- D. The licensee is authorized to collect leak test samples in accordance with the procedures described in the licensee's application dated July 8, 1980, for analysis by Nuclear Sources and Services, Inc. Alternatively, leak test samples may be collected and/or analyzed by other persons specifically authorized by the Commission or an Agreement State to perform such services.
15. The licensee shall conduct a physical inventory every six (6) months to account for all foils received and possessed under the license. The records of the inventories shall be maintained for two (2) years from the date of the inventory for inspection by the Commission, and shall include the quantities and kinds of licensed material, location of foils, and the date of the inventory.
16. Detector cells containing licensed material shall not be opened or the foil sources removed from the detector cell by the licensee.

MATERIALS LICENSE
SUPPLEMENTARY SHEET

License number

45-19479-01

Docket or Reference number

030-17756

Amendment No. 01

(cont'd)

CONDITIONS

17. Except as specifically provided otherwise by this license, the licensee shall possess and use licensed material described in Items 6, 7, and 8 of this license in accordance with statements, representations, and procedures contained in applications dated July 8, 1980 and June 25, 1985. The Nuclear Regulatory Commission's regulations shall govern the licensee's statements in applications or letters, unless the statements are more restrictive than the regulations.



FOR THE U.S. NUCLEAR REGULATORY COMMISSION

EARL G. WRIGHT

Date AUG 01 1985

By

*Earl G Wright*Region II, Nuclear Materials
Safety Section101 Marietta Street, Suite 2900
Atlanta, GA 30323

CONVERSATION RECORD

TIME

DATE

TYPE

☐ VISIT

☐ CONFERENCE

☐ TELEPHONE

ROUTING

NAME SYMBOL INT

Location of Visit/Conference

☐ INCOMING

☒ OUTGOING

NAME OF PERSON(S) CONTACTED OR IN CONTACT WITH YOU

ORGANIZATION (Office Dept. Bureau)

TELEPHONE NO

DAVID BUSHMAN

FAIRFAX CTY. GOVT
LORTON, VA.

8 (202)
550-9740

SUBJECT

RENEWAL OF Lic. #45-19479-01

SUMMARY

CALL TO EXPLAIN THAT RENEWAL OF LICENSE
WILL BE BASED ON OLD & NEW APPLICATIONS
SINCE SOME AREAS CLEARER (AND STILL SAME) IN
OLD APPLICATION:

- 1) MANUFACTURER & MODEL # OF DETECTOR CEU
- 2) LEAK TEST KIT STILL BEING USED
- 3) FOR DISPOSAL THEY WOULD RETURN TO
MANUFACTURER

ACTION REQUIRED

NONE

NAME OF PERSON DOCUMENTING CONVERSATION

SIGNATURE

DATE

C. CONNELL

C. Connell

7-29-85

ACTION TAKEN

SIGNATURE

TITLE

DATE

50271-101

© 1951 G - 361-526 (7227)

CONVERSATION RECORD

OPTIONAL FORM 271 (12-76)
DEPARTMENT OF DEFENSE