

Honeywell

July 30, 1985

Mr. Lester Tripp
NRC-Region I
631 Park Avenue
King of Prussia, PA 19406

Dear Mr. Tripp:

In our phone conversation of 7/25/85 you requested some additional information pertaining to Honeywell's recently submitted "Application For Material License."

The following should become a part of our application:

1. Monitoring devices will be supplied by:

R.S. Landauer, Jr. & Co.,
Glenwood Science Park
Glenwood, IL 60425

2. The Landauer B-1 is recommended for Am:Be sources and will be exchanged monthly.
3. The storage area for our Troxler gauge will be a padlocked cabinet in a normally unoccupied room in the Maintenance Department.
4. Access to this gauge will be restricted to the extent that only two employees will be issued keys to this cabinet:

K. Kittel - Facilities Engineer
M. Chunko - Radiation Safety Officer

If you require any additional information please contact me at 215-641-4328.

Yours truly,

Michael J. Chunko

Michael J. Chunko
Industrial Hygienist

8510280189 850911
REG1 LIC30
37-20846-01 PDR

"OFFICIAL RECORD COPY"

MJC/b

ML10

103802

8/2/85

7/25/85

TELEPHONE OR VERBAL CONVERSATION RECORD

TIME

☐ A.M.
☐ P.M.

☐ INCOMING CALL

☐ OUTGOING CALL

☐ VISIT

PERSON CALLING

J.M. Tripp

OFFICE/ADDRESS

R I

PHONE NUMBER

EXTENSION

PERSON CALLED

Mike Chunks

OFFICE/ADDRESS

Honeywell, Inc

PHONE NUMBER

EXTENSION

215-641-4328

CONVERSATION

SUBJECT

Material license application for Honeywell, Inc. DFFR/FAC/

SUMMARY

TELEPHONE DEFICIENCIES

- 1- Who will supply personnel monitoring devices?
- 2- How often will personnel monitoring devices be exchanged?
- 3- Will the gauge be locked & secured while in storage?
- 4- Who will keep the key to storage area?

REFERRED TO:

ACTION REQUESTED

Response to above questions

☐ ADVISE ME OF ACTION TAKEN.

INITIALS

DATE

ACTION TAKEN

Responses 2/B forward in written form by M. Chunks

INITIALS

DATE

"OFFICIAL RECORD COPY"

MLTB