



Edward Suter
Supervisor - Environmental Quality Control

Consolidation Coal Company
Eastern Region
450 Racetrack Road
Washington, Pennsylvania 15301
(412) 746-3400

September 26, 1985

John E. Glenn, Ph.D., Chief
Nuclear Materials Safety Section B
Division of Radiation Safety and Safeguards
United States Nuclear Regulatory Commission
Region I
631 Park Avenue
King of Prussia, Pennsylvania 19406

RE: License No. 37-15550-02
Inspection No. 030-13689/85-01
Docket No. 030-13689

Dear Sir:

Pursuant to the inspection conducted by Mr. M. A. Varela of your office and the subsequent issuance of a Notice of Violation, the following information is submitted to abate the violation:

The source rods were immediately locked in the closed position. The keys are being retained by the Radiation Safety Officer. The storage room where the gauges are located has limited access by employees only. A list of employees authorized to remove the gauges for use has been posted in the storage room. The room is locked at the end of the work day and the building is patrolled by a security guard during times when the office is normally closed.

If you should require any additional information in order to abate the Notice of Violation, please contact this office.

Sincerely,

A handwritten signature in cursive script that reads "Edward Suter".

Edward Suter

ES:akd

cc: M. W. Hunt

8510280319 851023
REG1 LIC30
37-15550-02 PDR

SURFACE NUCLEAR DENSITY GAUGES

ONLY THE FOLLOWING PERSONS ARE AUTHORIZED TO REMOVE THE 3411 B TROXLER
NUCLEAR DENSITY GAUGES FROM THIS STORAGE ROOM:

DENNIS FREDERICKS
DAN GABLER
REGE PETERNEL
BOB KOMOSINSKI
MIKE SADOWSKI
WILLIAM FERTALL
SCOTT RICE
ED SUTER

PRIOR TO REMOVAL, SEE ED SUTER FOR THE KEY TO THE SOURCE ROD LOCK, FILM
BADGE AND LOG BOOK.