



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION I
631 PARK AVENUE
KING OF PRUSSIA, PENNSYLVANIA 19406

June 17, 1985

DOCKET NO: 030-21271
CONTROL NO. 03903

Capaldi Bros. Corp.
ATTN: Michael A. Favicchio
Vice President
226 So. Main Street
Providence, RI 02901

Gentlemen:

This is in reference to your application dated May 24, 1985, for a byproduct material license. In order to continue our review, we need the following additional information:

1. If you intend to have Bernard L. Alderson specifically named on the license as an authorized user, then you need to submit a copy of his certificate of training in the safe use of Troxler's portable nuclear gauges. Please submit this documentation.
2. Please describe your procedures for securing the portable gauge from unauthorized access or removal by unauthorized users while being stored in your "stock room" that you designated as the permanent storage area. Procedures should be described on how you will secure the gauge while it is out of its carrying case for re-charging.
3. Since you will remove the source rod from the portable gauge(s) for routine cleaning and re-lubrication of the shaft and bearings, it will be necessary for you to have available a G-M survey meter for measuring radiation levels around the area that the source rod is placed to ensure that such levels do not exceed the limits set forth in 10 CFR 20.105 (enclosed). Accordingly, please specify the name of the manufacturer and model number of the radiation survey meter you will use to accomplish this. Specify the name of the company who calibrates this instrument, and the NRC or Agreement State license number that authorizes this company to do this.

We will continue our review of your application upon receipt of the above information. Please reply in duplicate, referencing Control No. 03903.

8508130334 850726
REG1 LIC30
38-20858-01 PDR

"OFFICIAL RECORD COPY"

ML10

Sincerely,

Original Signed By:

John E. Glenn

John E. Glenn, Ph.D., Chief
Nuclear Materials Safety Section B
Division of Radiation Safety and
Safeguards

Enclosure: 10 CFR 20

RI:DRSS

Varela

6/17/85

RI:DRSS

Glenn

6/17/85

OFFICIAL RECORD COPY