

**MATERIALS LICENSE  
SUPPLEMENTARY SHEET**

License number

48-17480-01

Docket or Reference number

030-12825

Amendment No. 03

Wisconsin Industrial Testing, Inc.  
4250 N. 126th Street  
Milwaukee, WI 53005

In accordance with application dated July 3, 1985, License Number 48-17480-01 is amended as follows:

Items 6., 7., 8. and 9. are amended to add:

6. Byproduct, source,  
and/or special nuclear  
material

7. Chemical and/or  
physical form

8. Maximum amount that  
licensee may possess  
at any one time  
under this license

F. Cesium-137

F. Sealed source(s)  
(Gamma Industries  
Model VD-HP)

F. No single source  
to exceed 225  
millicuries

9. Authorized Use

F. For use in a Gamma Industries Master Minder Model 2 exposure device as a control device for x-ray pipe crawlers.

Condition 17. is amended to read:

17. Except as specifically provided otherwise by this license, the licensee shall possess and use licensed material described in Items 6, 7, and 8 of this license in accordance with statements, representations, and procedures contained in applications dated September 27, 1982 and July 3, 1985 (with attachments); and letter dated July 13, 1983. The Nuclear Regulatory Commission's regulations shall govern the licensee's statements in applications or letters, unless the statements are more restrictive than the regulations.

Conditions 18., 19., 20. and 21. are added:

18. Sealed sources containing licensed material shall not be opened or removed from their respective source holders by the licensee.

19. When performing tests at temporary job sites, the authorized user shall not leave the Gamma Industries Master Minder Model 2 exposure device unattended. Upon completion of tests the device shall be locked in the licensee's vehicle or a secure building to prevent unauthorized use, loss or theft.

8508130401 850705  
REG3 LIC30  
48-17480-01 PDR

COPY

211  
HL 30 P.V.

**MATERIALS LICENSE  
SUPPLEMENTARY SHEET**

License number

48-17480-01

Docket or Reference number

030-12825

Amendment No. 03

20. Repair of the Gamma Industries Master Minder Model 2 exposure devices containing licensed material and replacement and disposal of sealed sources containing licensed material used in these devices shall be performed only by the device manufacturer or by other persons specifically authorized by the Commission of an Agreement State to perform such services.
21. The licensee shall have available and follow the step-by-step operating procedures and the step-by-step inspection and maintenance procedures contained in the Gamma Industries Master Minder Model 2 instruction manual.



For the U.S. Nuclear Regulatory Commission

Original Signed

By J.R. Madera

Materials Licensing Section, Region III

Date July 5, 1985

COPY