



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION I
631 PARK AVENUE
KING OF PRUSSIA, PENNSYLVANIA 19406

JUL 25 1985

Docket No. 030-21253
Control No. 103763

Ken Plummer, Jr., Inc.
ATTN: Noel Schooley
Safety Officer
527 Lincoln Way East
Chambersburg, PA 17201

SUBJECT: APPLICATION FOR MATERIAL LICENSE DATED APRIL 29 1985,
AND OUR REQUEST FOR INFORMATION IN A TELEPHONE CONVERSATION
MAY 24, 1985

Gentlemen:

This concerns the subject application for a material license and to the telephone conversation between yourself and Mr. Miller of my staff on May 24, 1985.

We have not received the requested information. Specifically, you have not confirmed that you will have a survey meter on hand to adequately assess the radiation levels that will be present when the source rod is removed. In addition, please specify how your survey meter will be calibrated.

Also confirm that portable gauges will be transported in accordance with DOT regulations.

You are hereby notified that unless within thirty (30) days from the date of this notice we receive the additional information requested, we will consider that you have abandoned your application. This action is without prejudice to the resubmission of an application.

Sincerely,

Lawrence F. Friedman

for John E. Glenn, Ph.D., Chief
Nuclear Materials Safety Section B
Division of Radiation Safety
and Safeguards

8510290453 851004
REG1 LIC30
37-20842-01 PDR

AUG 19 1985