

JAN 12 1982

FCMLB-PCJ  
030-19371  
(00358)

BioClinical Group, Inc.  
ATTN: Dr. Ernest V. Groman  
Senior Scientist  
767 B. Concord Avenue  
Cambridge, Massachusetts 02138

Gentlemen:

This is in reference to your application for NRC license dated December 11, 1981. In order to continue our review, we need the following additional information:

1. Information concerning item 12, "Personnel Monitoring Devices", is missing from your application. Please provide the name of the supplier of this service. Please confirm that individuals who iodinate protein will wear finger TLD's to monitor exposure to the hands as well as film badges to monitor exposure to the whole body.
2. Please provide the name of the commercial waste disposal service you will use. (Item 14 of the application form).
3. In regard to your procedure for safely opening packages, please confirm that the radiation level on the outside of the package will be measured with a survey meter, the outside of the inner lead pig will be wiped and measured for contamination, the packing material will be surveyed for contamination, and records of these surveys will be maintained.
4. In association with your procedure for "Filter Changing in Radiation Lab Hood Stacks", please confirm that item 4.5 (air sampling) will be performed during iodination operations.
5. Item 1.3 under Procedure on page 36 of your manual makes reference to animal tissue and excretion which indicates you will be conducting animal studies using radioactive material. If this is so, please describe your animal facilities and provide your instructions to caretakers.

RI:DETI  
Glenn/dmg  
1/8/82

RI:DETI  
for Jermann  
1/8/82

OFFICIAL RECORD COPY

8510290629 850916  
REG1 LIC30  
20-20526-01 PDR

ML10

BioClinical Group, Inc.

2

JAN 12 1982

We will continue our review of your application upon receipt of this information. Please reply in duplicate and refer to Control No. 00358.

Sincerely,

Original Signed By:

Phillip C. Jerman

Phillip C. Jerman  
Material Licensing Branch  
Division of Fuel Cycle and  
Material Safety

Enclosure: Regulatory Guide 8.20

OFFICIAL RECORD COPY