

MATERIALS LICENSE
SUPPLEMENTARY SHEET

License number

06-15099-02

Docket or Reference number

030-19430

Amendment No. 03

Canberra Clinical Labs., Inc.
75 North Mountain Road
New Britain, Connecticut 06052

In accordance with letter dated April 26, 1985, License Number 06-15099-02 is amended as follows:

Items. 6., 7., 8. and 9. are amended to read:

- | | | |
|---|----------------------------------|--|
| 6. Byproduct, source, and/or special nuclear material | 7. Chemical and/or physical form | 8. Maximum amount that licensee may possess at any one time under this license |
| A. Iodine 125 | A. Bound to nonvolatile agents. | A. 10 millicuries |

9. Authorized use

- A. In vitro clinical assays.

Condition 14. is amended to read:

14. Except as specifically provided otherwise by this license, the licensee shall possess and use licensed material described in Items 6, 7, and 8 of this license in accordance with statements, representations, and procedures contained in application dated October 6, 1982, and letters dated December 17, 1982, excluding item 7, November 15, 1983, July 16, 1984 and April 26, 1985. The Nuclear Regulatory Commission's regulations shall govern the licensee's statements in applications or letters, unless the statements are more restrictive than the regulations.

Condition 15. is added:

15. Radioactive wastes containing microcuries amounts of iodine-125 may be disposed to the ordinary trash after being held for decay for a minimum of five (5) half-lives. Prior to disposal, these wastes must be monitored in accordance with the procedures described in the licensee's letter dated November 15, 1983. The survey conducted prior to disposal must confirm that the radioactivity of the wastes cannot be distinguished from background.

For the U.S. Nuclear Regulatory Commission

Original Signed By:
Laurence F. Friedman, Ph.D.

Date

JUL 24 1985

By

Nuclear Materials Safety and
Safeguards Branch, Region I
King of Prussia, Pennsylvania 19406

8508140028 850724
REG1 LIC30
06-15099-02 PDR

"OFFICIAL RECORD COPY"

ML10