

ENCLOSURE

NOTICE OF VIOLATION

Florida Power and Light Company
St. Lucie

Docket Nos. 50-335, 50-389
License Nos. DPR-67, NPF-16

The following violation was identified during an inspection conducted on June 24-28, 1985. The Severity Level was assigned in accordance with the NRC Enforcement Policy (10 CFR Part 2, Appendix C).

10 CFR 50, Appendix B, Criterion II, requires that activities affecting quality be accomplished under suitably controlled conditions, including suitable environmental conditions for accomplishing the activity. The licensee's Quality Assurance Manual, QP 12.1, paragraph 5.5, requires that measuring and test equipment (M&TE) shall be calibrated in environments which will not adversely affect their accuracy and that the following factors shall be considered: temperature, humidity, vibration, radio frequency interference, electromagnetic interference, background radiation, dust, cleanliness, and fumes.

Contrary to the above, adequate measures have not been established to assure suitable environmental conditions for the calibration of M&TE in that specific limits defining acceptable environmental conditions have not been established for the calibration of M&TE in the Instrumentation and Control (I&C) and Mechanical shops.

This is a Severity Level IV violation (Supplement I).

Pursuant to 10 CFR 2.201, you are required to submit to this office within 30 days of the date of this Notice, a written statement or explanation in reply, including: (1) admission or denial of the alleged violation; (2) the reasons for the violation if admitted; (3) the corrective steps which have been taken and the results achieved; (4) corrective steps which will be taken to avoid further violations; and (5) the date when full compliance will be achieved.

Security or safeguards information should be submitted as an enclosure to facilitate withholding it from public disclosure as required by 10 CFR 2.790(d) or 10 CFR 73.21.

Date: July 29, 1985

8508140045 850729
PDR ADOCK 05000335
Q PDR