

SEP 26 1985

Docket No. : 030-17558
License No.: 50-19401-01
Control No.: 70234

Ted Forsi & Associates, Inc.
810 East Ninth Avenue, Suite 200
Anchorage, Alaska 99501

Attention: Mr. Nils Lindholm
Radiation Protection Officer

Gentlemen:

Enclosed is the NRC license renewal which you requested.

Please review the enclosed document carefully and be sure that you understand all conditions. If there are any errors or questions, please notify us so that we can provide appropriate corrections and answers.

Please note that License Condition 14, requires that any maintenance of the gauges involving dismantling, removal of source holder(s), etc., must not be performed by the user and must only be performed by the manufacturer of the device. If you wish to conduct such activities, you should request a license amendment. You should submit the information as described on page 9 of the enclosed guide.

It is NRC policy that either film badges or thermoluminescent dosimeters (TLD) should be worn when portable gauges are used. Refer to Item 10.1, Personnel Monitoring Equipment, on pages 7 and 8 of the enclosed guide. As TLDs may be exchanged on a quarterly basis, you may prefer to use this type of personnel dosimetry.

As no training or experience information was submitted for Mr. Ron Towne, he was not specifically listed on the license. However, if he has completed the manufacturer's training course and a copy of his training certificate is on file in your office, then he is an authorized user. Refer to License Condition 12.

Any future correspondence relating to your license should specifically reference your license and docket numbers to expedite your inquiry.

Sincerely,

8510310121 850926
REGS LIC30
50-19401-01 PDR

Beth A. Riedlinger
Health Physicist (Licensing)

Nuclear Materials Safety Section

OFFICE	RV	B.A.R.				
SURNAME	Riedlinger:fr	9/26/85				
DATE	Enclosures	Attachment A				
		Amendment No. 3 to License No. 50-19401-01				
		Portable gauge licensing guide				