

SEP 23 1985

Docket No. : 030-01221  
License No.: 04-01935-03  
Control No.: 70233

V. A. Medical Center  
2615 East Clinton Avenue  
Fresno, California 93703

Attention: Mr. J.J. Mason  
Director

Gentlemen:

Enclosed is the NRC license amendment which you requested.

Please review the enclosed document carefully and be sure that you understand all conditions. If there are any errors or questions, please notify us so that we can provide appropriate corrections and answers.

The xenon-133 traps used in the nuclear medicine areas should be tested weekly by checking the effluent from a patient study by using a plastic collection bag and a scintillation camera. The results should be logged and compared with previous tests to determine if the trapping efficiency has begun to fail.

The ventilation systems in the xenon use and storage areas should be checked every 6 months to assure that concentrations in restricted and unrestricted areas do not exceed the limits stated in 10 CFR 20.103 and 10 CFR 20.106.

Any future correspondence relating to your license should specifically reference your license and docket numbers to expedite your inquiry.

Sincerely,

8510310131 850923  
REG5 LIC30  
04-01935-03 PDR

Beth A. Riedlinger  
Health Physicist (Licensing)  
Nuclear Materials Safety Section

Enclosure: Attachment A  
Amendment No. 36 to License No. 04-01935-03

OFFICE	RV <i>BAR</i>						
SURNAME	Riedlinger:fr						
DATE	9/23/85						