

MAV 01 1985

City of Appleton
Wastewater Division
ATTN: Peter C. Wachs
Superintendent
59 Weimar Court
Appleton, WI 54915

Gentlemen:

We are in receipt of your letter dated March 27, 1985 requesting an amendment to NRC License Number 48-18213-01 for use of a Ramsey Engineering belt scale gauging device containing cesium-137. The information submitted, however, is not sufficiently detailed for us to complete our health and safety review.

We need to know the specific model number of the gauge and source holder, the purpose for which the gauge will be used, the names and qualifications of the individuals responsible for supervising the use of the device, a description of your facilities and the environmental conditions to which the gauge will be subjected and certain information on the radiation safety program that you have developed for use of the device.

To assist you in this effort we have enclosed a licensing guide for use of sealed sources in nonportable gauging devices. You should address each of the items described on pages 4 through 13. You may use regular 8½" x 11" paper to document your response.

Please reply in duplicate within 30 days and reference Control No. 78638. If you have questions or require additional clarification on the above, you may contact me on (312) 790-5625.

Be advised that Ramsey Engineering is the parent company of Texas Nuclear Corporation. The gauging device which you are requesting is probably registered with the NRC under the name of the latter. You should contact Ramsey Engineering for assistance.

Sincerely,

Original Signed By
B. J. Holt
Materials Licensing Section

Enclosure: Nonportable Gauge
Licensing Guide

RIII

Holt/lid
04/25/85

8508140257 850801
REG3 LIC30
48-18213-01 PDR