

J. H. MAXYMILLIAN, INC.

General Contracting

OFFICE: 1801 EAST STREET • PITTSFIELD, MA 01201 • PHONE: (413) 499-3050

June 24, 1985

MS 16
P8

U.S. Nuclear Regulatory Commission
Region I
631 Park Avenue
King of Prussia, PA 19406

Attention: John E. Glenn, Ph.D.

Reference: Mail Control No. 103874

Gentlemen:

The following information is provided in response to your request of June 11, 1985, concerning the review of our application for a byproduct material license:

1. We will be using a Landauer type personnel monitoring device from the Landauer Tech OPS Co. The B1 type badge that we will be using has an exchange frequency of one month.
2. The supplier of our gauge has indicated to us that it will not be necessary to remove the source rod during cleaning.

If you have any further questions concerning our application, please feel free to contact me. We have provided this letter in duplicate, as requested.

Very truly yours,

J. H. MAXYMILLIAN, INC.

Anthony J. Simonelli
Anthony J. Simonelli
Area Superintendent

AJS/ds

Certified Mail P620528213

8508140278 850724
REG1 LIC30
20-20853-01 PDR

RECEIVED-REG
JUN 26 PM 1:39
9505 JUN 26 PM 1:39

03874

"OFFICIAL RECORD COPY"

ML10

JUN 26 1985