

ENCLOSURE

NOTICE OF VIOLATION

Materials Testing Laboratory of Virginia, Inc.

License No. 45-17151-01

The following violation was identified during an inspection conducted on July 11, 1985. The Severity Level was assigned in accordance with the NRC Enforcement Policy (10 CFR Part 2, Appendix C).

10 CFR 34.24 requires a licensee to calibrate each radiation survey meter used to conduct physical radiation surveys at intervals not to exceed three months.

Contrary to the above, survey meter no. 832, a meter used to conduct physical radiation surveys on February 25 - 26, 1985, was last calibrated in November 21, 1984, an interval in excess of three months.

This is a Severity Level V violation (Supplement VI).

Pursuant to 10 CFR 2.201, you are required to submit to this office within 30 days of the date of this Notice, a written statement or explanation in reply, including: (1) admission or denial of the alleged violations; (2) the reasons for the violations if admitted; (3) the corrective steps which have been taken and the results achieved; (4) corrective steps which will be taken to avoid further violations; and (5) the date when full compliance will be achieved.

Security or safeguards information should be submitted as an enclosure to facilitate withholding it from public disclosure as required by 10 CFR 2.790(d) or 10 CFR 73.21.

Date: JUL 23 1985