

AUG 01 1985

Docket No. 30-14580

License No. 06-02354-02

John B. Pierce Foundation of Connecticut, Inc.  
ATTN: Dr. James Douglas, Ph.D.  
Associate Fellow  
290 Congress Avenue  
New Haven, Connecticut 06519

Gentlemen:

Subject: Inspection No. 30-14580/85-01

This refers to your letter dated June 27, 1985, in response to our letter dated May 28, 1985.

Thank you for informing us of the corrective and preventive actions documented in your letter. These actions will be examined during a future inspection of your licensed program.

With regard to your question about the requirement for calibrating survey meters which are not appropriate for the isotopes currently being used, you are advised that you are required to maintain a calibrated survey which is sufficiently sensitive to the radiations emitted by the isotopes being used. There is no need to maintain or calibrate survey meters which are incapable of measuring these radiations.

Your cooperation with us is appreciated.

Sincerely,

Original Signed By:  
John D. Kinneman

John D. Kinneman, Chief  
Nuclear Materials Section A  
Nuclear Materials Safety & Safeguards Branch  
Division of Radiation Safety & Safeguards

cc:  
Public Document Room (PDR)  
Nuclear Safety Information Center (NSIC)  
State of Connecticut

8508150223 850801  
REG1 LIC30  
06-02354-02 PDR

OFFICIAL RECORD COPY

RL PIERCE - 0001.0.0  
07/26/85

RETURN ORIGINAL TO  
REGION I


IF-07

AUG 01 1985

John B. Pierce Foundation  
of Connecticut, Inc.


2

bcc:  
Region I Docket Room (w/concurrences)

 R1:DRSS

Costello/djh

7/24/85

 R1:DRSS

Kinneman

7/24/85

OFFICIAL RECORD COPY

RL PIERCE - 0002.0.0  
07/26/85