

ENCLOSURE

NOTICE OF VIOLATION

International Circuit Technology

License No. 45-23485-01

The following violation was identified during an inspection conducted on July 15, 1985. The Severity Level was assigned in accordance with the NRC Enforcement Policy (10 CFR Part 2, Appendix C).

License Condition 16 requires that licensed material be used in accordance with statements and procedures contained in your application dated October 2, 1984. One statement in your application requires you to have a G-M survey meter available for use.

Contrary to the above, it was determined during the inspection that no survey meter is available for use at your facility.

This is a Severity Level IV violation (Supplement VI).

Pursuant to 10 CFR 2.201, you are required to submit to this office within 30 days of the date of this Notice, a written statement or explanation in reply, including: (1) admission or denial of the alleged violations; (2) the reasons for the violations if admitted; (3) the corrective steps which have been taken and the results achieved; (4) corrective steps which will be taken to avoid further violations; and (5) the date when full compliance will be achieved.

Security or safeguards information should be submitted as an enclosure to facilitate withholding it from public disclosure as required by 10 CFR 2.790(d) or 10 CFR 73.21.

Date: AUG 02 1985

8508150416 850802
REG2 LIC30
45-23485-01 PDR