

APPENDIX A  
NOTICE OF VIOLATION

Stanley Walter Siegler, M.D.  
Robert J. Gorczyca, M.D.  
Toms River, New Jersey 08753

Docket No. 030-02507  
License No. 29-07892-01

As a result of the inspection conducted on June 24, 1985, September 13, 1985 and October 1, 1985, and in accordance with the NRC Enforcement Policy (10 CFR 2, Appendix C), the following violations were identified:

Condition 15 of License No. 29-07892-01 requires that licensed material be possessed and used in accordance with statements, representations and procedures contained in application dated April 26, 1983.

1. Item 9, page 6 of this application requires that survey meters be calibrated at least annually and following repair.

Contrary to the above, as of August 7, 1985, one survey meter had not been calibrated since April 25, 1984 and the other not since September 24, 1982 periods which exceed one year.

2. Block 10 of the application requires that dose calibrators be calibrated in accordance with procedures in Appendix D, Section 2, of Regulatory Guide 10.8.

Item E. of Appendix D, Section 2, requires that in the test for linearity, time intervals of 6, 24, 30 and 48 hours be used after the initial assay, use assigned correction factors to determine predicted activities for required time intervals and plot the measured net activity for each interval versus the calculated activity.

Contrary to the above, as of October 1, 1985, the time interval did not extend to 48 hours after the initial assay, the assigned correction factors were not used, and the results were not plotted and evaluated as required.

These are Severity Level IV violations. (Supplement VI)

Pursuant to the provisions of 10 CFR 2.201, Stanley Walter Siegler, M.D. and Robert J. Gorczyca, M.D. are hereby required to submit to this office within thirty days of the date of the letter which transmitted this Notice, a written statement or explanation in reply, including: (1) the corrective steps which have been taken and the results achieved; (2) corrective steps which will be taken to avoid further violations; and (3) the date when full compliance will be achieved. Where good cause is shown, consideration will be given to extending this response time.

OFFICIAL RECORD COPY

DL SEIGLER - 0004.0.0  
10/24/85

8511110027 851031  
REG1 LIC30  
29-07892-01 PDR