

GENESEE MEMORIAL HOSPITAL, INC.

702 BALLENGER HIGHWAY

FLINT, MICHIGAN 48502 (313) 239-1481

July 24, 1985

Mr. D. J. Sreniawski  
United States Nuclear Regulatory Commission  
Region III  
799 Roosevelt Road  
Glen Ellyn, Illinois 60137

Re: License No. 21-15131-01

Dear Mr. Sreniawski:

This letter is in response to the non-compliance items identified in the Hospital survey of May 21, 1985, by Mr. W. P. Reichhold.


Item #1 - L. R. Irish, M.D., will be identified as the Radiation Safety Officer subsequent to his application receiving favorable approval by the Nuclear Regulatory Commission. Dr. Irish has stated his application is in process.

Item #2 - The Isotopes Committee met this date, July 23, 1985, and will meet quarterly thereafter. The Nuclear Medicine Department and the Hospital Administrative Officer for the Department will track these meetings to insure that they meet quarterly as scheduled.

Item #3 - The Hospital Nuclear Physicist, Mr. Jeffery Colvin, will visit the Hospital on Friday, July 26, 1985, and provide the Hospital technician with the necessary formula to comply to measurement standards required by the N.R.C. for wipe tests conducted in Nuclear Medicine.

I would like to thank Mr. Reichhold for his telephone call regarding the survey findings and for his assistance in interpreting the leak test records requirements.

Sincerely,

  
Homer W. Read  
President

HWR:nb

8508160054 850812  
REG3 LIC30  
21-15131-01 PDR

JUL 26 1985