

MATERIALS LICENSE

Amendment No. 11

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 36, 39, 40, and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations, and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

301745

Licensee		In accordance with the letter dated June 19, 1996	
1. St. Joseph's Community Hospital of West Bend, Inc.		3. License Number 48-16117-01 is amended in its entirety as follows:	
2. 551 S. Silverbrook Drive West Bend, WI 53095		4. Expiration Date January 31, 2002	
		5. Docket or Reference No. 030-10451	
6. Byproduct, Source, and/or Special Nuclear Material	7. Chemical and/or Physical Form	8. Maximum Amount that Licensee May Possess at Any One Time Under This License	
A. Any byproduct material identified in 10 CFR 35.100	A. Any radiopharmaceutical identified in 10 CFR 35.100	A. As needed	
B. Any byproduct material identified in 10 CFR 35.200	B. Any radiopharmaceutical identified in 10 CFR 35.200	B. As needed	
C. Any byproduct material identified in 10 CFR 35.300	C. Any radiopharmaceutical identified in 10 CFR 35.300	C. As needed (not to exceed 1 curie of I-131)	
D. Any byproduct material identified in 10 CFR 31.11	D. Prepackaged Kits	D. As needed	

Authorized Use:

- A. Medical use described in 10 CFR 35.100.
- B. Medical use described in 10 CFR 35.200.
- C. Medical use described in 10 CFR 35.300.
- D. In vitro studies.

9609250173 960905
PDR ADDCK 03010451
C PDR

COPY

MATERIALS LICENSE
SUPPLEMENTARY SHEETLicense Number
48-16117-01Docket or Reference Number
030-10451

Amendment No. 11

CONDITIONS

10. Licensed material shall be used only at the licensee's facilities located at 551 S. Silverbrook Drive, West Bend, Wisconsin, 1700 W. Paradise, West Bend, Wisconsin, and 205 Valley Avenue, West Bend, Wisconsin.

11. Radiation Safety Officer: Roland C. Brown, M.D.

12. Licensed material listed in Item 6 above is only authorized for use by, or under the supervision of, the following individuals for the materials and uses indicated:

Authorized UsersMaterial and Use

Roland C. Brown, M.D. 10 CFR 35.100, 35.200, 35.300 and 31.11

William J. Pier, M.D. 10 CFR 35.100, 35.200, 35.300 and 31.11

John G. Fink, M.D. 10 CFR 35.100, 35.200, 35.300 (excluding I-131 for thyroid carcinoma) and 31.11

13. The licensee is authorized to transport licensed material only in accordance with the provisions of 10 CFR Part 71, "Packaging and Transportation of Radioactive Material."


14. Except as specifically provided otherwise in this license, the licensee shall conduct its program in accordance with the statements, representations, and procedures contained in the documents, including any enclosures, listed below, except for minor changes in the medical use radiation safety procedures as provided in 10 CFR 35.31. The Nuclear Regulatory Commission's regulations shall govern unless the statements, representations, and procedures in the licensee's application and correspondence are more restrictive than the regulations.

- A. Application dated July 24, 1991;
- B. Letters received October 17, 1991 and February 20, 1996 (with enclosures);
- C. Letters dated December 18, 1991, January 16, 1992, and June 19, 1996; and
- D. Close-out survey results dated August 13, 1996.

FOR THE U.S. NUCLEAR REGULATORY COMMISSION

Date 9/5/96

By


Nuclear Materials Licensing Branch, Region III

COPY

S2

(FOR LFMS USE)
INFORMATION FROM LTS

BETWEEN:

License Fee Management Branch, ARM
and
Regional Licensing Sections

Program Code: 02120
Status Code: 0
Fee Category: 7C
Exp. Date: 20020131
Fee Comments:
Decom Fin Assur Req'd: N

LICENSE FEE TRANSMITTAL

A. REGION

1. APPLICATION ATTACHED

Applicant/Licensee: ST. JOSEPH'S COMM.HOSP.OF WEST BEND
Received Date: 960819
Docket No: 3010451
Control No.: 301745
License No.: 48-16117-01
Action Type: Amendment

2. FEE ATTACHED

Amount: ~~XXXXXX~~
Check No.: ~~XXXXXX~~

* ADDL INFO
301521-S2

3. COMMENTS

Signed D. Hersey
Date 8-23-96

B. LICENSE FEE MANAGEMENT BRANCH (Check when milestone of fee is entered)

FEE NOT REQUIRED

1. Fee Category and Amount:

2. Correct Fee Paid. Application may be processed for:

Amendment ☒
Renewal ☐
License ☐

3. OTHER

Signed SC
Date 8/27/96

1996 AUG 26 AM 11:19

SEP 03 1996

Log	Aug 12 III
Remitter	
Check No.	
Amount	
Fee Category	7C
Type of Fee	AmD
Date Check Rec'd	
Date Completed	8/27/96
By:	SC

CLOSEOUT SURVEY

St. Joseph's Community Hospital --
Facility: WEST BEND CLINIC **Performed By:** Michele Uecker, CNMT
Patient Room No.: Lab Stress Room **Date:** August 13, 1996

SURVEY RESULTS

Area Surveyed	mR/hr.	Wipe DPM/100CM ²	Comments
BACKGROUND	0.02 mR/hr	238 DPM	
PREP AREA	0.01 mR/hr	161 DPM	
PATIENT INJ AREA	0.01 mR/hr	228 DPM	
STORAGE AREA # 1	0.02 mR/hr	319 DPM	
STORAGE AREA # 2	0.02 mR/hr	314 DPM	
SINK	0.02 mR/hr	301 DPM	

ACTION LEVELS: 0.1 MR/HR SURVEYS AND 2000 DPM/100CM² FOR WIPES

***SEE DIAGRAM FOR AREA REFERENCE**

Instrument used for survey: Victoreen 190 **Cal. Date** 11/20/95
Instrument used for wipe analysis: Victoreen 190 **Cal. Date** 11/20/95

Serial Number of Victoreen 190 -- 1201

☒ **Room released for general use**
☐ **Room restricted**

Comments: All areas surveyed less than action levels.

RSO:

RC Brown, MD **Surveyor:** Michele Uecker

Close-Out Survey
FEE NOT REQUIRED
ADD'L info - 301521


RECEIVED

AUG 19 1996

REGION III

301745
AUG 19 1996

WEST BEND CLINIC CLOSEOUT SURVEY DIAGRAM
ST. JOSEPH'S COMMUNITY HOSPITAL



1. BACKGROUND
2. PREP AREA
3. PATIENT INJECTION AREA
4. STORAGE AREA #1
5. STORAGE AREA #2
6. SINK

SEP 16 1996

Roland Brown, M.D.
Radiation Safety Officer
St. Joseph's Community Hospital
of West Bend, Inc.
551 S. Silverbrook Drive
West Bend, WI 53015

Dear Dr. Brown:

Enclosed is Amendment No. 11 to your NRC Material License No. 48-16117-01 in accordance with your request.

Please review the enclosed document carefully and be sure that you understand all conditions. If there are any errors or questions, please notify the U.S. Nuclear Regulatory Commission, Region III office at (630) 829-9887 so that we can provide appropriate corrections and answers.

Based upon the close-out survey results dated April 13, 1996, we have authorized the release of the former nuclear medicine area at the West Bend Clinic for unrestricted use.

Please be advised that your license expires at the end of the day, in the month, and year stated in the license. Unless your license has been terminated, you must conduct your program involving byproduct materials in accordance with the conditions of your NRC license, representations made in your license application, and NRC regulations. In particular, note that you must:

1. Operate in accordance with NRC regulations 10 CFR Part 19, "Notices, Instructions and Reports to Workers; Inspections," 10 CFR Part 20, "Standards for Protection Against Radiation," and other applicable regulations.
2. Notify NRC, in writing, within 30 days:
 - a. When an authorized user or Radiation Safety Officer permanently discontinues performance of duties under the license or has a name change; or
 - b. When the licensee's mailing address changes (no fee is required if the location of byproduct material remains the same).

3. In accordance with 10 CFR 30.36(b) and/or license condition, notify NRC, promptly, in writing, and request termination of the license when you decide to terminate all activities involving materials authorized under the license.
4. Request and obtain a license amendment before you:
 - a. Receive or use byproduct material for a clinical procedure permitted under Part 35 but not permitted by your license issued pursuant to this Part;
 - b. Permit anyone, except individuals described in 10 CFR 35.13(b), to work as an authorized user under the license;
 - c. Change Radiation Safety Officers;
 - d. Order byproduct material in excess of the amount, or radionuclide, or form different than authorized on the license;
 - e. Add or change the areas of use or address or addresses of use identified in the license application or on the license; or
 - f. Change ownership of your organization.
5. Submit a complete renewal application with proper fee or termination request at least 30 days before the expiration date of your license. You will receive a reminder notice approximately 90 days before the expiration date. Possession of byproduct material after your license expires is a violation of NRC regulations. A license will not normally be renewed, except on a case-by-case basis, in instances where licensed material has never been possessed or used.

In addition, please note that NRC Form 313 requires the applicant, by his/her signature, to verify that the applicant understands that all statements contained in the application are true and correct to the best of the applicant's knowledge. The signatory for the application should be the licensee or certifying official rather than a consultant.

You will be periodically inspected by NRC. Failure to conduct your program in accordance with NRC regulations, license conditions, and representations made in your license application and supplemental correspondence with NRC will result in enforcement action against you. This could include issuance of a notice of violation, or imposition of a civil penalty, or an order suspending, modifying or revoking your license as specified in the General Policy and Procedures for NRC Enforcement Actions. Since serious consequences to employees and the public can result from failure to comply with NRC requirements,

R. Brown

-3-

prompt and vigorous enforcement action will be taken when dealing with licensees who do not achieve the necessary meticulous attention to detail and the high standard of compliance which NRC expects of its licensees.

Sincerely,

Original Signed By
Michael F. Weber
Nuclear Materials Licensing Branch

License No.: 48-16117-01
Docket No.: 030-10451

Enclosure: Amendment No. 11

DOCUMENT NAME: M:\03010451.CL6

To receive a copy of this document, indicate in the box: "C" = Copy without attachment/enclosure "E" = Copy with attachment/enclosure "N" = No copy

OFFICE	DNMS/RIII	E							
NAME	MWEBER:jaw <i>mtw</i>								
DATE	09/5/96								

OFFICIAL RECORD COPY