

From: Kevin G. Null (KGN)
To: DAB
Date: Friday, September 23, 1994 8:30 am
Subject: Daybrerak irradiator

Doug,

Regarding the Daybreak irradiator we've discussed - After searching for more "clues" in the license file, I've discovered that the the device was authorized in Feb. of 1991. Both a Sr-90 and a Cm-244 source were authorized to be used in the unit. Upon further investigation, I found product infomation in the file on the model 801 which describes the use of both a beta and alpha source. This is interesting in light of the fact that the registry sheet only refers to us of a Sr-90 source. I thought you would be interested in this so I will FAX the product info. to you for your review.

I am sending a letter to the licensee in response to their 3/21/94 letter pertaining to removal of the Sr-90 source from the unit. I am using your comments that you E-mailed me.

I am confused with the apparent differences between the product info and the registry sheet for the model 801 as it pertains to sources approved for the device. Please let me know what becomes of this. Once again, thanks very much for your assistance in this matter.

Kevin Null