

NOTE TO NEXT LICENSE REVIEWER FOR LICENSE NO. 06-17253-01 (VICTOR J. BORTOLOTT, PH.D.)

1. Appropriate program code for this license is 03122, measuring systems analytical instruments. Authorized use (Condition 9) must be written appropriately.
2. Licensee has requested additional sources to install into a new device. This device will require a source and device review by HQ. This new device should also be clearly delineated in Condition 9. The licensee's two current devices were grandfathered under 32.210 and do not require any further review.
3. Have the licensee either commit to obtaining a custom source and device review prior to selling a device to another licensee who does not possess a broad scope license or commit to having his customer obtain the source and device review prior to receiving the device. In addition, require the licensee to maintain copies of appropriate records to demonstrate that his clients are properly licensed to use his device.
4. Have the licensee commit or incorporate into the license that the licensee is not authorized to perform any type of service of his equipment at the customer's location. If wishes to work on his equipment at a client location, it must be done under the client license.
5. Licensee will require to possess a survey instrument to comply with Condition No. 15 of his license.
6. The Sr-90 source (Condition 6C) used by the licensee to leak test his other sources is not exempt quantity or generally licensed. At the time of the last inspection (92-001), the licensee was unable to demonstrate to the NRC that it was anything but a specifically licensed.

Questions on the above? Talk to either myself or Sandra Kimberly first regarding this licensee.

Duncan White
12/22/92

NOTE TO NEXT LICENSE REVIEWER FOR LICENSE NO. 06-17253-01 (VICTOR J. BORTOLOTT, PH.D.)

1. Appropriate program code for this license is 03122, measuring systems analytical instruments. Authorized use (Condition 9) must be written appropriately.
2. Licensee has requested additional sources to install into a new device. This device will require a source and device review by HQ. This new device should also be clearly delineated in Condition 9. The licensee's two current devices were grandfathered under 32.210 and do not require any further review.
3. Have the licensee either commit to obtaining a custom source and device review prior to selling a device to another licensee who does not possess a broad scope license or commit to having his customer obtain the source and device review prior to receiving the device. In addition, require the licensee to maintain copies of appropriate records to demonstrate that his clients are properly licensed to use his device.
4. Have the licensee commit or incorporate into the license that the licensee is not authorized to perform any type of service of his equipment at the customer's location. If wishes to work on his equipment at a client location, it must be done under the client license.
5. Licensee will require to possess a survey instrument to comply with Condition No. 15 of his license.
6. The Sr-90 source (Condition 6C) used by the licensee to leak test his other sources is not exempt quantity or generally licensed. At the time of the last inspection (92-001), the licensee was unable to demonstrate to the NRC that it was anything but a specifically licensed.

Questions on the above? Talk to either myself or Sandra Kimberly first regarding this licensee.

Duncan White
12/22/92