

Note to File

2/15/95

Scott Markwell / Sleepsake

3:15 p.m.

telecon w/

(703) 989-5738

Told him:

- prototype testing of a similar device/chamber assembly is acceptable - but make an evaluation to show the test support the conclusion
 - prototype testing standards available from previous SD submittals - British Standards, etc.
- On slightly changed design and new model series, send in addendum to tie the new/added information to the original submittal. No additional review will be performed for the new reg'ing.

R Tang