

APPENDIX A

NOTICE OF VIOLATION

Kuakini Medical Center  
347 N. Kuakini Street  
Honolulu, Hawaii

License No. 53-17797-01

As a result of the inspection conducted July 11-12, 1985 and in accordance with the NRC Enforcement Policy, 10 CFR Part 2, Appendix C, the following violations were identified:

License Condition 20 requires that except as specifically provided otherwise by the license, the licensee shall possess and use licensed material in accordance with statements, representations, and procedures contained, in part, in the application dated August 26, 1983.

- A. Item 12 of the application states, in part, that ancillary personnel (i.e. clerical, nursing, security, maintenance and housekeeping) will be given annual instruction pertinent to their duties in the form of videotapes and color slides, and that all personnel who work with and handle radioactive materials will be given annual retraining.
- (1) Contrary to the above requirement, at the time of the inspection, a number of ancillary personnel had not received any refresher training as required on an annual basis. For several individuals the lack of refresher training has exceeded two years.
- (2) Contrary to the above requirement, at the time of the inspection, the Bone Mineral Technician in the Osteoporosis Laboratory had not received any refresher training as required on an annual basis. For this individual the lack of refresher training has exceeded three years.

The above items constitute a Severity Level IV Violation (Supplement VI).

- B. License Condition 17 states, in part, that each sealed source containing licensed material, other than hydrogen 3, with a half-life greater than 30 days and in any form other than gas shall be tested for leakage and/or contamination at intervals not to exceed six months.

Contrary to the above requirement, leak tests were conducted late as follows:

<u>Source</u>	<u>S.N.</u>	<u>Date of Leak Tests</u>	<u>Months Between Leak Tests</u>
12 mCi Am-241	77302	6-27-83 and 7-17-84	12
211 Ci Cs-137	NEN-356	12-2-83 and 7-17-84	17
1.5 Ci Gd-153		5-15-84 and 3-11-85	10

This is a repetitive violation.


This is a Severity Level IV Violation (Supplement VI).

8508190347 850813  
REG5 LIC30  
53-17797-01 PDR

Pursuant to the provisions of 10 CFR 2.201, Kuakini Medical Center is hereby required to submit to this office within thirty days of the date of this Notice, a written statement of explanation in reply, including: (1) the corrective steps which have been taken and the results achieved; (2) corrective steps which will be taken to avoid further violations; and (3) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

AUG 13 1985

Dated

  
R. D. Thomas, Chief  
Nuclear Materials Safety Section