



INSTRUMENT CONTROL SERVICE 4008 Dauphin Island Parkway • Mobile, Ala. 36605 • 205/479-6535

June 24, 1985

United States Nuclear Regulatory Commission
Region 1
631 Park Avenue
King of Prussia, Pennsylvania 19406

Attn: John D. Kinneman, Chief
Nuclear Materials Safety Section A,
Division of Radiation Safety and
Safeguards

Re: Docket #30-20004
License #01-20628-01

Gentlemen:

In reply to your letter of June 13, 1985, in which ICS was found to have a security level IV violation during inspection #30-20004/85-01 the following steps have been taken to rectify the problem:

1. A second Victoreen survey meter has been added to the site inventory. This unit is a model 493 and carries serial number 4361.
2. Calibration of the survey meters has been contracted with:
Radiation Service Organization, Inc.
5204 Minnick Road
Laurel, Maryland 20707

Both meters have been calibrated and a schedule of every three months has been established for future calibrations. An operational survey meter will be available on site at all times.

3. In respect to the violations noted we believe that full compliance to our licensing requirements was reestablished on June 4, 1985, when the second survey meter was placed in service.

It is our hope that this letter will serve to clear the violations. If there are any further steps required, please do not hesitate to contact me.

Sincerely,

INSTRUMENT CONTROL SERVICE

Richard McConnell
Richard McConnell
Manager of Services

RMC/lmw

8508200062 850731
REG1 LIC30
01-20628-01 PDR