



*Anheuser-Busch, Inc.*  
ONE OF THE ANHEUSER-BUSCH COMPANIES

Monday, September 23, 1996

DCD

Yes - *[initials]*  
10 CFR 31.5(c)(5).

*General Licensee*

Nuclear Regulatory Commission  
Region III  
861 Warrenville Road  
Lisle, IL 60532-4351

**Subject: Notification of Failed Leak Test of Electron Capture Device**

Mr. James Cameron:

Title 10 of the Code of Federal Regulations, Section 31.5 (c)(5), requires owners of radiation devices to notify the Nuclear Regulatory Commission (NRC) of possible failures of these devices.

On August 26, 1996, wipe tests were conducted for sources at the St. Louis complex. A failed test result was received for a Gas Chromatograph nickel 63 source, serial number 620. This unit was out of service prior to the wipe test. It will remain out of service as we proceed with disposal of the unit with the vendor, Perkin-Elmer.

If you have any questions or need additional information, please contact me at (314) 577-3584.

Sincerely,

*[Signature: Chad Fisher]*

Chad Fisher  
Environmental Engineer  
St. Louis Brewery

010045

9611010218 960923  
PDR I&E  
MISC PDR

Anheuser-Busch, Inc.  
One Busch Place  
St. Louis, MO U.S.A. 63118-1852  
Telex 447 117 ANBUSCH STL

1/0  
IE-72

SEP 27 1996

*I&E MISC*