

SAFETY INSPECTION

DMB COPY

1. LICENSEE

OAKLAWN HOSPITAL
200 NORTH MADISON ST.
MARSHALL, MI 49068

2. REGIONAL OFFICE

U.S. NUCLEAR REGULATORY COMMISSION
REGION III
799 ROOSEVELT ROAD
GLEN ELLYN, ILLINOIS 60137

3. DOCKET NUMBER(S)

30-12139

4. LICENSE NUMBER(S)

21-17068-01

5. DATE OF INSPECTION

8-13-85

Licensee:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission's (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews, with personnel, and observations by the inspector. The findings as a result of this inspection are as follows:

- ☐ 1. Within the scope of this inspection, no violations were observed.
- ☐ 2. The inspector also verified the steps you have taken to correct the violations identified during the last inspection. We have no further questions or those actions at this time.
- ☒ 3. During this inspection certain of your activities, as checked below, were in violation of NRC requirements.
THIS IS A NOTICE OF VIOLATION which is required to be posted in accordance with 10 CFR 19.11.
- ☐ A. _____ was not properly posted to indicate the presence of a _____ 10 CFR 20.203(b), (c), (d), (e) or 34.42.
- ☐ B. Containers located in _____ were not properly labeled to indicate the presence of radioactive material. 10 CFR 20.203(f)(1), or (f)(2).
- ☒ C. LEAK TESTS _____ of sealed sources were not performed at the proper frequencies. 10 CFR 35.14(e)(1) License Condition Number _____.
- ☐ D. Records of _____ were not properly maintained. 10 CFR _____ or License Condition Number _____.
- ☐ E. Documents were not properly posted or otherwise made available. 10 CFR 19.11.
- ☐ F. Reports or notifications of _____ were not made in accordance with 10 CFR _____ or License Condition Number _____.
- ☒ H. DOSE CALIBRATOR LINEARITY TEST HAS NOT BEEN DONE QUARTERLY AS REQUIRED BY LICENSE CONDITION 16, WHICH REFERENCES 6/10/81 APPLICATION.
- ☒ I. RADIATION SURVEY METER HAS NOT BEEN CALIBRATED ANNUALLY AS REQUIRED BY LICENSE CONDITION 16, WHICH REFERENCES 6/10/81 APPLICATION.
- ☒ J. INSTRUCTIONS ON THE SAFE USE OF LICENSED MATERIAL WAS NOT GIVEN TO EMPLOYEES ON ANNUAL BASIS AS REQUIRED BY LICENSE CONDITION 16, WHICH REFERENCES LETTER DATED 2/15/82.
- ☐ K. _____

8508280077 850813

REG3 LIC30

21-17068-01

PDR

I hereby state that within 30 days the actions described in this statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201. No further response will be submitted unless required by the NRC.

Violations identified in the items checked above.

James P. Kelly, Asst. Admin.
SIGNATURE - LICENSEE

8/13/85
DATE

S. R. Lesuk
SIGNATURE - NRC INSPECTOR

8/13/85
DATE

01 IE07