

Clayton Environmental Consultants, Inc.

Raritan Center, 160 Fieldcrest Ave., Edison, New Jersey 08837, Telephone (201) 225-6040

April 12, 1985

Mr. John E. Glenn, Chief
Nuclear Material Safety Section B
Division of Radiation Safety and Safeguards
United States Regulatory Commission
Region I
631 Park Avenue
King of Prussia, Pennsylvania 19406

Re: Docket No. 30-19440
License No. 29-20578-01
Inspection No. 030-19440/85-01

RECEIVED-REGION I
1985 APR 15 PM 4:00

Dear Mr. Glenn:

In response to your letter dated March 4, 1985, which was received on March 18, 1985, Clayton Environmental Consultants has initiated action to comply with your request. We respectfully submit the following information.

We would like to point out a minor error which we feel should also be corrected in the records of the following:

- Public Document Room (PDR)
- Nuclear Safety Information Center (NSIC)
- State of New Jersey

During 1983 and 1984, licensed material in question (AID - Portable Gas Chromatograph detector equipped with tritium source.) was used at our Southfield, Michigan corporate headquarters and not at Chicago facility. We do not have any location in Chicago.

As a result of the inspection conducted on February 15, 1985 by Ms. J. Joustra at your office, the violation was issued in accordance with 10 CFR 30.34 (c) and condition 10 of license No. 29-20578-01. To correct the cited violation Clayton has adopted the following policy:

- I: The source in question will not leave 160 Fieldcrest Avenue, Edison, New Jersey until our license has been amended. All Clayton employees have been instructed.
- II: In order to avoid further violation, we will be officially transferring the source to our Southfield, Michigan location. We will amend both licenses with said changes.

8508280325 850524
REG1 LIC30
29-20578-01 PDR


Clayton Environmental Consultants

III: Until said date when licenses have been amended through the Local NRC Region, the source will not be shipped. Any future shipping of the source will be accomplished only after proper documentation in accordance with NRC and DOT requirement has been completed.

Please provide us with all necessary forms needed to properly implement this plan. Your guidance in this matter is requested and appreciated.

If you require further information, please contact this office.

Very truly yours,

A handwritten signature in dark ink, appearing to read 'Kirit H. Vora', written over a horizontal line.

Kirit H. Vora, Manager
New Jersey Office and Laboratory

KHV:ss