

MATERIALS LICENSE  
SUPPLEMENTARY SHEET

License number

45-01121-04

Docket or Reference number

030-00514

Amendment No. 14

Department of the Navy  
Naval Hospital  
Portsmouth, Virginia 23708

In accordance with letter dated July 2, 1985, License Number 45-01121-04 is amended as follows:

Condition 23. is amended to read:

23. Pursuant to Section 20.105(a) of Title 10, Chapter 1, Code of Federal Regulations, Part 20, "Standards for Protection Against Radiation," and in reliance of statements, procedures and representations made by the licensee in his letter with enclosures dated July 2, 1985 the following maximum radiation levels are hereby authorized in the following unrestricted areas:

Maximum Radiation Level

45 milliroentgens per hour

Unrestricted Area

Stretcher Patient Waiting Area

FOR THE U. S. NUCLEAR REGULATORY COMMISSION

DATE Aug. 21, 1985

BY Francis A. H. Mary  
Material Licensing Branch  
Division of Fuel Cycle and  
Material Safety  
Washington, D. C. 20555

8508280382 850821  
NMSS LIC30  
45-01121-04 PDR

11/1/85  
Hand Copy to Region II

2