



DEPARTMENT OF THE NAVY

NAVAL HOSPITAL

PORTSMOUTH, VIRGINIA 23708

IN REPLY REFER TO:

6470

7162

2 July 1985

From: Commanding Officer, Naval Hospital, Portsmouth, Virginia 23708  
To: Radioisotope License Branch, Division of Fuel Cycle and Material  
Safety, Office for Material Safety and Safeguards, U.S. Nuclear  
Regulatory Commission, Washington, DC 20555  
Via: Commander, Naval Medical Command (NAVMED-21),  
Washington, DC 20372

Subj: REQUEST FOR AMENDMENT TO U.S. NUCLEAR REGULATORY COMMISSION BYPRODUCT  
MATERIAL LICENSE NO. 45-01121-04

Ref: (a) Byproduct Material License No. 45-01121-04

Encl: (1) Teletherapy Radiation Survey Report

1. In accordance with reference (a), enclosure (1) is submitted.
2. It is requested that subject license be amended to authorize a maximum radiation level of 45 mR/hr at location (A) in enclosure (1) - a corridor and stretcher patient waiting area along the southwest wall. Enclosure (1) demonstrates compliance with 10 CFR 20.105(a) for this location. Our present license authorizes a value of 38 mR/hr based on measurements made with the old Cobalt-60 source.

G. J. ADLER  
By direction

Copy to:  
Commander, Naval Medical Command,  
Mid-Atlantic Region,  
Norfolk, VA 23508  
NRC Regional Office of Inspection and  
Enforcement, 101 Marietta St., NW,  
Suite 2900, Atlanta, GA 30323

8508280387 850821  
NMSS LIC30  
45-01121-04 PDR