

MAR 21 1985

Docket No. 030-22298
Control No. 18710

Delaware Technical and Community College
Stanton Campus
ATTN: Jerry N. Blancato
Biochemistry Instructor
333 Shipley Street
Wilmington, Delaware 19801

Gentlemen:

This is in reference to your application dated February 21, 1985 for a byproduct material license. In order to continue our review, we need the following additional information:

1. In your application you did not describe your laboratory instructions for working with radioactive materials. It is recommended that your procedures include, but not be limited to, the following topics:
 - a. mouth pipeting
 - b. protective clothing, gloves, etc.
 - c. storage of food, drink, or personnel effects
 - d. personnel dosimetry
2. In your application, you did not describe your procedures for opening packages containing radioactive material [10 CFR 20.205(d)]. Please specify your procedure for surveying the radiation and contamination levels on those packages. In addition, please state who will assume the responsibility for monitoring these packages.

We will continue our review upon receipt of this information. Please reply in duplicate to my attention at the Region I office and refer to Mail Control No. 18710.

Sincerely,

Original Signed By:
Jack Davis

Jack Davis
Nuclear Materials Safety Section A
Division of Radiation Safety
and Safeguards

Enclosure: 10 CFR Part 20

RI:DRSS

Davis/mlb

3/4/85

OFFICIAL RECORD COPY

DELAWARE TECH & COM COL/D/LTR - 0001.0.0
03/20/85

8508280418 850813
REG1 LIC30
07-23516-01 PDR

ML10