

APPENDIX A

NOTICE OF VIOLATION

St. Joseph's Hospital and Medical Center
Paterson, New Jersey 07503

Docket No. 30-366
License No. 29-10191-01

As a result of the inspection conducted on February 4, 1985, and in accordance with the NRC Enforcement Policy (10 CFR 2, Appendix C), the following violation was identified:

Condition 14.A of License No. 29-10191-01 requires that each teletherapy source be tested for leakage at intervals not to exceed six (6) months.

Contrary to the above, as of February 4, 1985, the teletherapy source had not been leak tested since July 7, 1983.

This is a Severity Level IV Violation (Supplement IV).

Pursuant to the provisions of 10 CFR 2.201, St. Joseph's Hospital and Medical Center is hereby required to submit to this office within thirty days of the date of the letter which transmitted this Notice, a written statement or explanation in reply, including: (1) the corrective steps which have been taken and the results achieved; (2) corrective steps which will be taken to avoid further violations; and (3) the date when full compliance will be achieved. Where good cause is shown, consideration will be given to extending this response time.

8509030379 850827
REG1 LIC30
29-10191-01 PDR