

MATERIAL TESTING LABORATORIES OF VIRGINIA, INC.

2711 NEVADA AVENUE • NORFOLK, VIRGINIA 23513

(804) 855-1971

85 AUG 23 47:48



August 20, 1985

30-12279

U. S. Nuclear Regulatory Commission
Region II
101 Marietta Street, N.W.
Atlanta, Georgia 30323

Attn: Mr. Kenneth P. Barr, Chief, Nuclear Materials Safety
and Safeguards Branch

Subject: Report No. 45-17151-01/85-01


Gentlemen:

As required by your notice of violation dated July 23, 1985 our response is:

- (1) We admit the violation.
- (2) Although we keep a status board on survey meter calibrations, the man who maintained the board was in the hospital at the time of the infraction and the man filling in for him failed to send the meter off.
- (3) The meter was sent off for calibration three (3) days after the infraction and calibrations have been kept up to date since that time or meters have been removed from service.
- (4) Since receipt of amendment #8 to our license we have been performing our own calibrations. Our policy is to calibrate all meters at one time a week before they are due. In addition personnel are to note the calibration due date on the Daily Use Report.
- (5) We feel that we are now in full compliance.

Very truly yours,

Material Testing Laboratories of Va., Inc.


Ralph H. Johnson, Jr.
President

RHJ/ds

8509040023 850820
REG2 LIC30
45-17151-01 PDR

1/0 IEC7