

SEALED SOURCE FILES

L&R:IB:WOM

MAY 5 1961

Tracerlab Inc.
1601 Treadle Road
Waltham 54, Massachusetts

Attention: Mr. William Karp

Gentlemen:

Reference is made to your letter dated April 25, 1961, regarding your Model CD-1 radiography source.

The source appears to be acceptable for licensing purposes. Any of your customers submitting applications for this source should identify it by model number.

Sealed sources used outside exposure devices should contain the following information in addition to that shown in your drawing No. B-994144: isotope, quantity and date of measurement of the quantity.

Very truly yours,

William O. Miller
Senior Licensing Reviewer
Isotopes Branch
Division of Licensing and
Regulation

9702240397 610504
PDR RC * PDR
SSD

L&R:IB

WOMiller/sk

5/4/61

9702240397

SEALED SOURCES			
Manufacturer TRACKLAB INCORPORATED Waltham 54, Massachusetts		Distributor TRACKLAB INCORPORATED Waltham, Massachusetts	
Use	Isotope(s)	Model #	
Radiography Source	Cobalt 60	CD-1	

Description: Ballare welded stainless steel capsule measuring 0.25 x 0.561 inches. Capsule has "eye" at one end which may be used to secure source for, or for handling equipment.

May 4, 1961