

SAFETY INSPECTION

Page 1 of 1

1. LICENSEE St. Luke's Episcopal Hospital Caribbean Nuclear Pharmacy P.O. Box 984 Ponce, Puerto Rico 00733-0984		2. REGIONAL OFFICE REGION II U S NUCLEAR REGULATORY COMMISSION 101 MARIETTA STREET NW SUITE 2900 ATLANTA GA 30323	
3. DOCKET NUMBER(S) 030-34161	4. LICENSE NUMBER(S) 52-16061-02MD	5. DATE OF INSPECTION 1/27/97	



LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The findings as a result of this inspection are as follows:

- ☐ 1. Within the scope of this inspection, no violations were observed.
- ☐ 2. The inspector also verified the steps you have taken to correct the violations identified during the last inspection. We have no further questions on those actions at this time.
- ☒ 3. During this inspection certain of your activities, as described below or attached, were in violation of NRC requirements. This form is a **NOTICE OF VIOLATION**, which is required to be posted in accordance with 10 CFR 19.11.
- ☐ A. _____ was not properly posted to indicate the presence of a _____ 10 CFR 20.203(b),(c),(d),(e) or 34.42.
- ☐ B. _____ of sealed sources were not performed at the proper frequencies. 10 CFR _____ or License Condition Number _____.
- ☐ C. Records of _____ were not properly maintained. 10 CFR _____ or License Condition Number _____.
- ☐ D. Documents were not properly posted or otherwise made available. 10 CFR 19.11.
- ☐ E. Reports or notification of _____ were not made in accordance with 10 CFR _____ or License Condition Number _____.
- ☒ F. DOSE CALIBRATION LINEARITY NOT PERFORMED WITHIN QUARTER 90 DAYS AS REQUIRED BY 10CFR27.12 LC-21

9702250034 970127
PDR ADOCK 03034161
C PDR

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified in the items checked above. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201. No further response will be submitted unless required by the NRC.

SIGNATURE -- LICENSEE 	DATE Jan/27/97	SIGNATURE -- NRC INSPECTOR 	DATE 1/27/97
---	-------------------	--	-----------------