

MATERIALS LICENSE
SUPPLEMENTARY SHEET

License Number

13-25966-01

Docket or Reference Number

030-30784

Amendment No. 02

302120

National-Standard Company
2401 N. Home Street
Mishawaka, IN 46545

In accordance with the letter received December 11, 1996, License Condition 13-25966-01 is hereby terminated.

TERMINATED

FOR THE U.S. NUCLEAR REGULATORY COMMISSION

Date 2/13/97

By Michael F. White
Nuclear Materials Licensing Branch, Region III

9702260059 970213
PDR ADOCK 03030784
C PDR

COPY 2 ^{ml} 30 SD

BETWEEN:

License Fee Management Branch, ARM
and
Regional Licensing Sections

(FOR LFMS USE)
INFORMATION FROM LTS

Program Code: 03120
Status Code: 0
Fee Category: 3P
Exp. Date: 20040131
Fee Comments:
Decon Fin Assur Req'd: N

S2
16

LICENSE FEE TRANSMITTAL

A. REGION

1. APPLICATION ATTACHED

Applicant/Licensee: NATIONAL-STANDARD COMPANY
Received Date: 961211
Docket No: 3030784
Control No.: 302120
License No.: 13-25966-01
Action Type: Termination

2. FEE ATTACHED

Amount:
Check No.:

3. COMMENTS

Signed D. Harney
Date 12-16-96

B. LICENSE FEE MANAGEMENT BRANCH (Check when milestone 03 is entered / ✓ /)

1. Fee Category and Amount: 3P

2. Correct Fee Paid ✓ Application may be processed for:
Amendment
Renewal
License

3. OTHER

Signed
Date 12/22/96

1996 DEC 19 PM 3:20

Log	Dec 6 1996
Remitter	
Check No.	
Fee Category	3P
Type of Fee	Amendment
Date Check Rec'd	
Date Completed	12/22/96
By	

National Standard Co.
2401 N. Home st.
Mishawaka, Indiana
46545

United States
Nuclear Regulatory Commission
Region III
801 Warrenville Road
Lisle, Illinois 60532 - 4351

To Whom It May Concern,


We would like to Cancel our Specific License #
13-25966-01 which is for two Strontium 90 density gauges.
These gauges will be under a General license according to
10CFR 11.5.

One of the gauges has been turned over to the
manufacturer. There License Number is 1663, issued by the
State of California. All surveys and reading of wipe tests
will be done by them. They have a specific license. We will
perform only the things that are allowed under the General
license.


The State of Indiana has been notified and we will
register the gauge in Indiana under a General license.

Please inform me if there is anything else that I need
to do.

Best regards,


Steve Hurt
Production Supervisor, RSO

P.S.- Sent duplicate letter to 101 Marietta Street NW, Suite
2900 Atlanta, Ga. 30323.


TERMINATION
FEE EXEMPT

RECEIVED
DEC 11 1996
REGION III

PM: 12-9-96

DEC 11 1996
302/20

FEB 13 1997

Steven L. Hurt
Production Supervisor
National-Standard Company
2401 N. Home Street
Mishawaka, IN 46545-4499

Dear Mr. Hurt:

Enclosed is Amendment No. 02 which terminates your NRC License Number 13-25966-01 in accordance with your request.

If you have any questions or require clarification on any of the information stated above, you may contact us at (630) 829-9887.

Sincerely,

Original Signed By
Michael F. Weber
Nuclear Materials Licensing Branch

License No.: 13-25966-01
Docket No.: 030-30784

Enclosure: Amendment No. 02

DOCUMENT NAME: M:\03030784.TR7

To receive a copy of this document, indicate in the box: "C" = Copy without attachment/enclosure "E" = Copy with attachment/enclosure "N" = No copy

OFFICE	DNMS/RIII	<input checked="checked" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
NAME	MWEBER:jaw	<input checked="checked" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
DATE	02/13/97	<input checked="checked" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>

OFFICIAL RECORD COPY

302120

NS

National-Standard

Nonwoven Metal Products

January 28, 1997

Michael F. Weber
Nuclear Materials Licensing Branch
Region III
801 Warrenville Road
Lisle, Illinois 60532

Dear Mr. Weber:

I am writing a reply to your Control Number 302120. We have installed the correct label on our gauge clearly stating all the information that is requested in 10 CFR 32.51 (a) (3). We will continue wipe tests and shutter on/off tests every six months.

Please be assured that all regulations in 10 CFR 31.5 © will be followed for our general license.

Sincerely,



Steve Hurt
Production Supervisor
RSO

SH:eh

RECEIVED
JAN 31 1997
REGION III

pm: 1-29-97

2401 N. Home Street • Mishawaka, Indiana 46545-4499 • Tel. 219/259-8505
Fax: 219/259-8507

JAN 31 1997

JAN 15 1997

Steven L. Hurt
Radiation Safety Officer
National Standard Company
2401 N. Home Street
Mishawaka, IN 46545

Dear Mr. Hurt:

We have reviewed your letter received December 11, 1996 requesting termination of NRC License No. 13-25966-01. This termination action is subject to reclassifying your gauge from a specifically licensed device to a generally licensed device. Before your request can be approved, you must provide the following:

1. Provide a commitment that the proper labels and instructions as required by 10 CFR 32.51(a)(3) will be attached to your device. For your reference, this regulation is included in Attachment 1 to this letter.
2. Provide a commitment that the regulations in 10 CFR 31.5(c) will be followed. For your reference, this regulation is included in Attachment 2 to this letter.

We will continue our review of your request upon receipt of this information. Please reply in duplicate, within 30 days, and refer to Control Number 302120.

If you have any questions or require clarification on any of the information stated above, you may contact us at (630) 829-9887.

Sincerely,

Original Signed By
Michael F. Weber
Nuclear Materials Licensing Branch

License No. 13-25966-01
Docket No. 030-30784

DOCUMENT NAME: M:\03030784.DF7

To receive a copy of this document, indicate in the box: "C" = Copy without enclosures "E" = Copy with enclosures "N" = No copy

OFFICE	DNMS/RIII	E							
NAME	MWEBER:jaw								
DATE	01/14/97								

OFFICIAL RECORD COPY

Attachment 1

10 CFR 32.51(a)(3)

(3) Each device [must] bear a durable, legible, clearly visible label or labels approved by the Commission which contain in a clearly identified and separate statement:

(i) Instructions and precautions necessary to assure safe installation, operation, and servicing of the device (documents such as operating and service manuals may be identified in the label and used to provide this information);

(ii) The requirements, or lack of requirements, for leak testing, or for testing any on-off mechanism and indicator, including the maximum time interval for such testing, and the identification of radioactive material by isotope, quantity of radioactivity, and date of determination of the quantity; and

(iii) The information called for in the following statement in the same or substantially similar form:

The receipt, possession, use, and transfer of this device Model _____¹, Serial No. _____¹, are subject to a general license or the equivalent and the regulations of the U.S. NRC or of a State with which the NRC has entered into an agreement for the exercise of regulatory authority. This label shall be maintained on the device in a legible condition. Removal of this label is prohibited.

CAUTION -- RADIOACTIVE MATERIAL

(Name of manufacturer, or initial transferor)¹

¹The model, serial number, and the name of the manufacturer, or initial transferor may be omitted from this label provided the information is elsewhere specified in labeling affixed to the device

Attachment 2

10 CFR 31.5(c)

(c) Any person who acquires, receives, possesses, uses or transfers byproduct material in a device pursuant to the general license in paragraph (a) of this section:

(1) Shall assure that all labels affixed to the device at the time of receipt and bearing a statement that removal of the label is prohibited are maintained thereon and shall comply with all instructions and precautions provided by such labels;

(2) Shall assure that the device is tested for leakage of radioactive material and proper operation of the on-off mechanism and indicator, if any, at no longer than six-month intervals or at such other intervals as are specified in the label; however:

(i) Devices containing only krypton need not be tested for leakage of radioactive material, and

(ii) Devices containing only tritium or not more than 100 microcuries of other beta and/or gamma emitting material or 10 microcuries of alpha emitting material and devices held in storage in the original shipping container prior to initial installation need not be tested for any purpose;

(3) Shall assure that the tests required by paragraph (c)(2) of this section and other testing, installation, servicing, and removal from installation involving the radioactive materials, its shielding or containment, are performed:

(i) In accordance with the instructions provided by the labels; or

(ii) By a person holding a specific license pursuant to parts 30 and 32 of this chapter or from an Agreement State to perform such activities;

(4) Shall maintain records showing compliance with the requirements of paragraphs (c)(2) and (c)(3) of this section. The records must show the results of tests. The records also must show the dates of performance of, and the names of persons performing, testing, installing, servicing, and removing from the installation radioactive material and its shielding or containment. The licensee shall retain these records as follows:

(i) Each record of a test for leakage of radioactive material required by paragraph (c)(2) of this section must be retained for three years after the next required leak test is performed or until the sealed source is transferred or disposed of.

(ii) Each record of a test of the on-off mechanism and indicator required by paragraph (c)(2) of this section must be retained for three years after the next required test of the on-off mechanism and indicator is performed or until the sealed source is transferred or disposed of.

(iii) Each record that is required by paragraph (c)(3) of this section must be retained for three years from the date of the recorded event or until the device is transferred or disposed of.

(5) Upon the occurrence of a failure of or damage to, or any indication of a possible failure of or damage to, the shielding of the radioactive material or the on-off mechanism or indicator, or upon the detection of 0.005 microcurie or more removable radioactive material, shall immediately suspend operation of the device until it has been repaired by the manufacturer or other person holding a specific license pursuant to parts 30 and 32 of this chapter or from an Agreement State to repair such devices, or disposed of by transfer to a person authorized by a specific license to receive the byproduct material contained in the device and, within 30 days, furnish to the Administrator of the appropriate Nuclear Regulatory Commission, Regional Office listed in appendix D of part 20 of this chapter, a report containing a brief description of the event and the remedial action taken;

(6) Shall not abandon the device containing byproduct material;

(7) Shall not export the device containing byproduct material except in accordance with part 110 of this chapter;

(8) Except as provided in paragraph (c)(9) of this section, shall transfer or dispose of the device containing byproduct material only by transfer to persons holding a specific license pursuant to parts 30 and 32 of this chapter or from an Agreement State to receive the device and within 30 days after transfer of a device to a specific licensee shall furnish to the Director of Nuclear Material Safety and Safeguards, U.S. Nuclear Regulatory Commission, Washington, DC 20555 a report containing identification of the device by manufacturer's name and model number and the name and address of the person receiving the device. No report is required if the device is transferred to the specific licensee in order to obtain a replacement device;

(9) Shall transfer the device to another general licensee only:

(i) Where the device remains in use at a particular location. In such case the transferor shall give the transferee a copy of this section and any safety documents identified in the label of the device and within 30 days of the transfer, report to the Director of Nuclear Material Safety and Safeguards, U.S. Nuclear Regulatory Commission, Washington, DC 20555, the manufacturer's name and model number of device transferred, the name and address of the transferee, and the name and/or position of an individual who may constitute a point of contact between the Commission and the transferee; or

(ii) Where the device is held in storage in the original shipping container at its intended location of use prior to initial use by a general licensee.

(10) Shall comply with the provisions of Sections 20.2201, and 20.2202 of this chapter for reporting radiation incidents, theft or loss of licensed material, but shall be exempt from the other requirements of parts 19, 20, and 21, of this chapter.

National Standard Co.
2401 N. Home st.
Mishawaka, Indiana
46545

United States
Nuclear Regulatory Commission
Region III
101 Marietta Street NW,
Suite 2900
Atlanta, Ga. 30323

To Whom It May Concern,

We would like to Cancel our Specific License #
13-25966-01 which is for two Strontium 90 density gauges.
These gauges will be under a General license according to
10CFR 31.5.

One of the gauges has been turned over to the
manufacturer. There License Number is 1663, issued by the
State of California. All surveys and reading of wipe tests
will be done by them. They have a specific license. We will
perform only the things that are allowed under the General
license.

The State of Indiana has been notified and we will
register the gauge in Indiana under a General license.

Please inform me if there is anything else that I need
to do.

Best regards



Steve Hurt
Production Supervisor, RSO

P.S.- Sent duplicate letter to 801 Warrenville Rd. Lisle,
Illinois 60532-4351

RECEIVED
DEC 23 1996
REGION III

302120
DEC 23 1996



UNITED STATES
NUCLEAR REGULATORY COMMISSION

REGION III
801 WARRENVILLE ROAD
LISLE, ILLINOIS 60532-4351

December 16, 1996

Steven L. Hurt
Radiation Safety Officer
National-Standard Company
2401 N. Home Street
Mishawaka, IN 46545

SUBJECT: ACKNOWLEDGEMENT OF CORRESPONDENCE
(Letter Received 12/11/96)

Dear Licensee:

In response to your request, we have completed the initial processing, which is an administrative review of your application for a(n):

☐ New License ☐ Amendment ☐ Renewal
☒ Termination ☐ Auth User (Amendment not required)
☐ Other _____

No administrative deficiencies were identified during this initial review. However, it should be noted that a technical review may identify omissions in the submitted information.

It appears that your request is routine (see 1-3 below, as applicable).

1. New and amendment actions are normally processed within 90 days, unless we find major deficiencies, or policy issues requiring central program office assistance.
2. Renewal actions are normally processed within 180 days, however, under timely filing (before expiration), you may continue to operate under your existing license.
3. Termination actions are normally processed within 90 days, unless confirmatory surveys following decontamination/decommissioning activities are involved.

A copy of your correspondence has been forwarded to our Licensing Fee and Debt Collection Branch (301/415-6097) for approval of the fee category and amount, if required.

If you have a compelling safety or business-related reason for requesting expedited review, please contact the Materials Licensing Branch at (630) 829-9887. We will try to complete your request as soon as practicable. Any correspondence about this request should reference the control number.

Nuclear Materials Support Branch

Mail Control No. 302120
License No. 13-25966-01

15-25966-01

National Standard
2401 N. Home Street
Michawaka, IN 46544

Director of Material Licensing
U. S. Nuclear Regulatory Commission
Washington, D.C. 20545

Gentlemen:

This is to advise you that we have returned the following radioactive source to Measurex Corporation, One Results Way, Cupertino, CA 95014:

One Strontium 90 sealed source, 25 mCi, s/n 10156, contained in an Ohmart BAL sensor

Measurex distributes its products containing radioactive material under California license GLR56-43. Measurex possesses radioactive material under license California 1663-43.

Very truly yours,



Signature

Radiation Safety Officer

Title

12/4/96

Date

CC: U. S. Nuclear Regulatory Commission
801 Warrenville Rd.
Lisle, IL 60532-4351