

Form AEC-313 (5-58)	ATOMIC ENERGY COMMISSION <b>APPLICATION FOR BYPRODUCT MATERIAL LICENSE</b>	Form approved. Budget Bureau No. 38-R027.4.
<p><b>INSTRUCTIONS.</b>—Complete Items 1 through 16 if this is an initial application. If application is for renewal of a license, complete only Items 1 through 7 and indicate new information or changes in the program as requested in Items 8 through 15. Use supplemental sheets where necessary. Item 16 must be completed on all applications. Mail three copies to: U.S. Atomic Energy Commission, Washington, D.C., 20545. Attention: Isotopes Branch, Division of Licensing and Regulation. Upon approval of this application, the applicant will receive an AEC Byproduct Material License. An AEC Byproduct Material License is issued in accordance with the general requirements contained in Title 10, Code of Federal Regulations, Part 30 and the Licensee is subject to Title 10, Code of Federal Regulations, Part 20.</p>		
1. (a) NAME AND STREET ADDRESS OF APPLICANT. (Institution, firm, hospital, person, etc.)  Molechem, Inc., Subsidiary of The Harshaw Chemical Company 1945 East 97th Street Cleveland, Ohio 44106		(b) STREET ADDRESS(ES) AT WHICH BYPRODUCT MATERIAL WILL BE USED. (If different from 1 (a).)  2240 Prospect Street Cleveland, Ohio
2. DEPARTMENT TO USE BYPRODUCT MATERIAL  Testing Department		3. PREVIOUS LICENSE NUMBER(S). (If this is an application for renewal of a license, please indicate and give number.)  29-6790-1
4. INDIVIDUAL USER(S). (Name and title of individual(s) who will use or directly supervise use of byproduct material. Give training and experience in Items 8 and 9.)  John J. Baum		5. RADIATION PROTECTION OFFICER (Name of person designated as radiation protection officer if other than individual user. Attach resume of his training and experience as in Items 8 and 9.)
6. (a) BYPRODUCT MATERIAL. (Elements and mass number of each.)  Americium <sup>241</sup>  Americium <sup>241</sup>  Bismuth <sup>207</sup>	(b) CHEMICAL AND/OR PHYSICAL FORM AND MAXIMUM NUMBER OF MILLICURIES OF EACH CHEMICAL AND/OR PHYSICAL FORM THAT YOU WILL POSSESS AT ANY ONE TIME. (If sealed source(s), also state name of manufacturer, model number, number of sources and maximum activity per source.)  Open flat plated source less than 1 microcurie (2 sources) (ORTEC)  Liquid source - less than 1 microcurie (Tracerlab)  Open flat plated source - less than 1 microcurie (ORTEC)	
7. DESCRIBE PURPOSE FOR WHICH BYPRODUCT MATERIAL WILL BE USED. (If byproduct material is for "human use," supplement A (Form AEC-313a) must be completed in lieu of this item. If byproduct material is in the form of a sealed source, include the make and model number of the storage container and/or device in which the source will be stored and/or used.)  Testing Molechem Radiation Detection Products		

61814

*[Handwritten signature]*

(Continued on reverse side)

## TRAINING AND EXPERIENCE OF EACH INDIVIDUAL NAMED IN ITEM 4 (Use supplemental sheets if necessary)

B. TYPE OF TRAINING	WHERE TRAINED	DURATION OF TRAINING	ON THE JOB (Circle answer)	FORMAL COURSE (Circle answer)
a. Principles and practices of radiation protection	General Electric Co. A.N.P.D.	5 Yrs.	<input checked="" type="radio"/> Yes <input type="radio"/> No	<input type="radio"/> Yes <input type="radio"/> No
b. Radioactivity measurement standardization and monitoring techniques and instruments	General Electric Co. A.N.P.D.	5 Yrs.	<input checked="" type="radio"/> Yes <input type="radio"/> No	<input type="radio"/> Yes <input type="radio"/> No
c. Mathematics and calculations basic to the use and measurement of radioactivity	General Electric Co. A.N.P.D.	5 Yrs.	<input checked="" type="radio"/> Yes <input type="radio"/> No	<input type="radio"/> Yes <input type="radio"/> No
d. Biological effects of radiation	General Electric Co. A.N.P.D.	5 Yrs.	<input checked="" type="radio"/> Yes <input type="radio"/> No	<input type="radio"/> Yes <input type="radio"/> No

## 9. EXPERIENCE WITH RADIATION. (Actual use of radioisotopes or equivalent experience.)

ISOTOPE	MAXIMUM AMOUNT	WHERE EXPERIENCE WAS GAINED	DURATION OF EXPERIENCE	TYPE OF USE
Am <sup>241</sup>	1 mc	General Electric ANP Dept. Cincinnati, Ohio	8 years	Nuclear instrument testing
Po <sup>210</sup>	1 mc	"		
Po-Be	10 curies	Molechem, Inc.		

## 10. RADIATION DETECTION INSTRUMENTS. (Use supplemental sheets if necessary.)

TYPE OF INSTRUMENTS (Include make and model number of each)	NUMBER AVAILABLE	RADIATION DETECTED	SENSITIVITY RANGE (mr/hr)	WINDOW THICKNESS (mg/cm <sup>2</sup> )	USE (Monitoring, surveying, measuring)
Molechem #30	2	L-B	.01 cpm	.02	surveying

## 11. METHOD, FREQUENCY, AND STANDARDS USED IN CALIBRATING INSTRUMENTS LISTED ABOVE.

Counting std. Am<sup>241</sup> plated source once each month.

## 12. FILM BADGES, DOSIMETERS, AND BIO-ASSAY PROCEDURES USED. (For film badges, specify method of calibrating and processing, or name of supplier.)

None

## INFORMATION TO BE SUBMITTED ON ADDITIONAL SHEETS

13. FACILITIES AND EQUIPMENT. Describe laboratory facilities and remote handling equipment, storage containers, shielding, fume hoods, etc. Explanatory sketch of facility is attached. (Circle answer) Yes ☒ No ☐ Chemical Lab Room with sealed source cabinet & fume hood
14. RADIATION PROTECTION PROGRAM. Describe the radiation protection program including control measures. If application covers sealed sources, submit leak testing procedures where applicable, name, training, and experience of person to perform leak tests, and arrangements for performing initial radiation survey, servicing, maintenance and repair of the source. Monthly Lab smears counted for activity.
15. WASTE DISPOSAL. If a commercial waste disposal service is employed, specify name of company. Otherwise, submit detailed description of methods which will be used for disposing of radioactive wastes and estimates of the type and amount of activity involved. No radioactive waste

## CERTIFICATE (This item must be completed by applicant)

16. THE APPLICANT AND ANY OFFICIAL EXECUTING THIS CERTIFICATE ON BEHALF OF THE APPLICANT NAMED IN ITEM 1, CERTIFY THAT THIS APPLICATION IS PREPARED IN CONFORMITY WITH TITLE 10, CODE OF FEDERAL REGULATIONS, PART 30, AND THAT ALL INFORMATION CONTAINED HEREIN, INCLUDING ANY SUPPLEMENTS ATTACHED HERETO, IS TRUE AND CORRECT TO THE BEST OF OUR KNOWLEDGE AND BELIEF.

Date 10 August 1964



Molechem, Inc.

Applicant named in item 1

John J. Baum, General Manager

Title of certifying official

WARNING.—18 U. S. C., Section 1001; Act of June 25, 1948; 62 Stat. 760, makes it a criminal offense to make a willfully false statement or representation to any department or agency of the United States or to any member within its jurisdiction.

DATE RECEIVED AUG 1 1964	EXPIRATION DATE Aug 31, 1966	ISSUE DATE	REVIEWER BWC
CONTROL NO. 61814	REFERENCE NO.	LICENSE NO. 34-6558-4 H66	AMENDMENT NO. NEW
ISOTOPE	FORM	POSSESSION LIMIT	
A. Am <sup>241</sup>	A. <del>Plated</del> Deposited sources	A. 2 sources not to exceed 1 uc per source	
B. Am <sup>241</sup>	B. Amg	B. 1 uc	
C. Bismuth 207	C. Deposited source	C. 1 source not to exceed 1 uc	
D.	D.	D.	
E.	E.	E.	
F.	F.	F.	
G.	G.	G.	
H.	H.	H.	

AUTHORIZED USE

A-C. Instrument calibration and testing

REMARKS: Letters, Phone Calls, Visits, Exemptions, Etc. (Use reverse side if necessary)

- ⑤ ②... facilities to be licensed at 2240 Prospect Street, Cleveland, Ohio
- ⑦ John J. Baum
- ⑪ A. Each deposited source of Am<sup>241</sup> shall be tested for leakage and/or contamination at intervals not to exceed three months. See the above...
- D. ...the license ...
- ①/4 ... exp. dtd. Aug 10, 1964

MAIL TO: Baum	DATE MAILED AUG 28 1964	REVIEWER RBC	DATE 8/28/64
------------------	----------------------------	-----------------	-----------------

B/17