

SEP 06 1985

Docket No. 30-20741

License No. 08-20702-01

Metropolitan Radiology Associates, CHTD
ATTN: Joel B. Bowers, M.D.
Partner
2139 Georgia Avenue, N.W., Suite 4B
Washington, D.C. 20060

Gentlemen:

Subject: Inspection Report No. 30-20741/85-01

This refers to your letter dated June 24, 1985, in response to our letter dated May 28, 1985.

Thank you for informing us of the corrective and preventive actions documented in your letter. Regarding Item No.2, you should assure that the method established for the evaluation of contamination wipes is sufficiently sensitive to detect 200 disintergrations per minute per 100 square centimeters.

Your actions will be examined during a future inspection of your licensed program.

Your cooperation with us is appreciated.

Sincerely,

Original Signed By:
John D. Kinneman

John D. Kinneman, Chief
Nuclear Materials Section A
Nuclear Materials Safety & Safeguards Branch
Division of Radiation Safety & Safeguards

cc:
Public Document Room (PDR)
Nuclear Safety Information Center (NSIC)
District of Columbia

bcc:
Region I Docket Room (w/concurrences)

OFFICIAL RECORD COPY

RL METROPOLITAN - 0001.0.0
08/22/85

8509100084 850906
REG1 LIC30
08-20702-01 PDR

RETURN ORIGINAL TO
REGION I

1E:07

Kinneman
1:DRSS
Darden/djh
8/28/85

Kinneman
1:DRSS
Kinneman
8/28/85

OFFICIAL RECORD COPY

RL METROPOLITAN - 0002.0.0
08/12/85