

APPENDIX A

NOTICE OF VIOLATION

GTE Sylvania Glass Incorporated
Wellsboro, Pennsylvania 16901

Docket No. 030-19358
License No. 37-20517-01

As a result of the inspection conducted on June 18, 1985, and in accordance with the NRC Enforcement Policy (10 CFR 2, Appendix C), the following violation was identified:

Condition 17 of License No. 37-20517-01 requires that licensed materials be possessed and used in accordance with the statements, representations and procedures contained in an application dated September 22, 1981.

Block 11.2 of this application requires that your radiation survey meter be calibrated every (6) months.

Contrary to the above, as of June 18, 1985, the GM survey meter (#3965M) used at your facility has not been calibrated since June 17, 1983, a period greater than six months.

This is a Severity Level IV violation (Supplement IV).

Pursuant to the provisions of 10 CFR 2.201, GTE Sylvania Glass Incorporated is hereby required to submit to this office within thirty days of the date of the letter which transmitted this Notice, a written statement or explanation in reply, including: (1) the corrective steps which have been taken and the results achieved; (2) corrective steps which will be taken to avoid further violations; and (3) the date when full compliance will be achieved. Where good cause is shown, consideration will be given to extending this response time.

OFFICIAL RECORD COPY

DL GTE - 0003.0.0
08/27/85

8509100094 850904
REG1 LIC30
37-20517-01 PDR